

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV022420\
 Data File : VV014697.D
 Acq On : 24 Feb 2020 14:20
 Operator : SY/MD
 Sample : L1607-12
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AP8

Quant Time: Feb 25 07:39:15 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM022120WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Feb 25 06:37:27 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	209158	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	200116	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	94596	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	23604	28.90	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	57.80%#
7) Chloroethane-d5	1.59	69	21981	37.13	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	74.26%
11) 1,1-Dichloroethene-d2	2.13	63	43097	28.30	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	56.60%#
21) 2-Butanone-d5	3.95	46	70080	92.29	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	92.29%
24) Chloroform-d	4.40	84	112596	42.10	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	84.20%
26) 1,2-Dichloroethane-d4	5.08	65	81321	47.56	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.12%
32) Benzene-d6	5.10	84	215933	40.77	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	81.54%
36) 1,2-Dichloropropane-d6	6.12	67	65547	42.44	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	84.88%
41) Toluene-d8	7.36	98	201607	39.52	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	79.04%#
43) trans-1,3-Dichloropropene-	7.66	79	36239	41.89	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	83.78%
47) 2-Hexanone-d5	8.14	63	54833	88.58	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	88.58%
57) 1,1,2,2-Tetrachloroethane-	10.25	84	95883	44.81	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	89.62%
64) 1,2-Dichlorobenzene-d4	11.67	152	86027	46.48	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	92.96%

Target Compounds

					Ovalue
12) 1,1-Dichloroethene	2.14	96	2087	3.273 ug/L #	1
20) cis-1,2-Dichloroethene	3.96	96	46769	32.568 ug/L	98
25) Chloroform	4.42	83	5588	2.156 ug/L	84
30) 1,1,1-Trichloroethane	4.66	97	10697	4.597 ug/L	97
34) Trichloroethene	5.96	95	264267	175.906 ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV022420\
Data File : VV014697.D
Acq On : 24 Feb 2020 14:20
Operator : SY/MD
Sample : L1607-12
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0AP8

Quant Time: Feb 25 07:39:15 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM022120WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Feb 25 06:37:27 2020
Response via : Initial Calibration

