

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV022421\
 Data File : VV020386.D
 Acq On : 23 Feb 2021 17:42
 Operator : SY/MD
 Sample : VV0224WBL01
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK704

Quant Time: Feb 24 11:36:33 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR022421WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Feb 24 03:03:54 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.622	114	297344	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.857	117	285196	5.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.255	152	136580	5.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	95121	5.04	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	100.80%
7) Chloroethane-d5	1.568	69	79071	5.23	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	104.60%
11) 1,1-Dichloroethene-d2	2.108	63	129142	3.46	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	69.20%
20) 2-Butanone-d5	3.902	46	176746	51.16	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	102.32%
24) Chloroform-d	4.355	84	173823	4.90	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.00%
26) 1,2-Dichloroethane-d4	5.037	65	88406	5.27	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	105.40%
32) Benzene-d6	5.053	84	354871	5.35	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	107.00%
36) 1,2-Dichloropropane-d6	6.072	67	103605	5.26	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	105.20%
41) Toluene-d8	7.320	98	320649	5.00	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	100.00%
43) trans-1,3-Dichloroprop...	7.628	79	35091	4.75	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	95.00%
46) 2-Hexanone-d5	8.095	63	122424	46.81	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	93.62%
56) 1,1,2,2-Tetrachloroeth...	10.223	84	64820	4.61	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	92.20%
66) 1,2-Dichlorobenzene-d4	11.632	152	129224	5.60	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	112.00%
Target Compounds						Qvalue
25) Chloroform	4.381	83	5023	0.10	ug/L	75
48) 2-Hexanone	8.153	43	15258	1.93	ug/L	# 82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV022421\
Data File : VV020386.D
Acq On : 23 Feb 2021 17:42
Operator : SY/MD
Sample : VV0224WBL01
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK704

Quant Time: Feb 24 11:36:33 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR022421WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Feb 24 03:03:54 2021
Response via : Initial Calibration

