

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV022424\
 Data File : VV034391.D
 Acq On : 24 Feb 2024 13:00
 Operator : SY/MD
 Sample : P1511-05DL 5X
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 DCND3DL

Quant Time: Feb 26 00:36:25 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR012524WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Feb 24 00:41:32 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	77189	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	70926	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.188	152	35433	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	33430	5.592	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	111.800%	
7) Chloroethane-d5	1.535	69	25433	4.857	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	97.200%	
11) 1,1-Dichloroethene-d2	2.060	65	14183	4.920	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	98.400%	
20) 2-Butanone-d5	3.799	46	43656	43.482	ug/L	-0.02
Spiked Amount	50.000	Range 40 - 130	Recovery	=	86.960%	
24) Chloroform-d	4.252	84	53958	4.967	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	99.400%	
26) 1,2-Dichloroethane-d4	4.944	65	26178	5.157	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	103.200%	
32) Benzene-d6	4.960	84	102916	4.836	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	96.800%	
36) 1,2-Dichloropropane-d6	5.989	67	28534	4.797	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	96.000%	
41) Toluene-d8	7.243	98	93739	4.793	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	95.800%	
43) trans-1,3-Dichloroprop...	7.561	79	6972	3.104	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	62.000%	
46) 2-Hexanone-d5	8.027	63	35661	44.476	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	88.960%	
56) 1,1,2,2-Tetrachloroeth...	10.156	84	17528	4.861	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	97.200%	
66) 1,2-Dichlorobenzene-d4	11.561	152	32073	4.977	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	99.600%	
Target Compounds						
14) Carbon disulfide	2.243	76	8126	0.466	ug/L #	93
16) Methylene chloride	2.452	84	8188	1.583	ug/L	97
17) Methyl tert-butyl Ether	2.706	73	20167	1.820	ug/L	98
18) trans-1,2-Dichloroethene	2.696	96	42051	7.656	ug/L	95
19) 1,1-Dichloroethane	3.117	63	1493	0.148	ug/L #	88
22) cis-1,2-Dichloroethene	3.818	96	7374	1.183	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV022424\
Data File : VV034391.D
Acq On : 24 Feb 2024 13:00
Operator : SY/MD
Sample : P1511-05DL 5X
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
DCND3DL

Quant Time: Feb 26 00:36:25 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR012524WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Feb 24 00:41:32 2024
Response via : Initial Calibration

