

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW022624\
 Data File : VW034415.D
 Acq On : 26 Feb 2024 18:19
 Operator : SY/MD
 Sample : P1535-01
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AB4

Quant Time: Feb 27 05:07:36 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR022624WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Feb 27 05:05:59 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	144649	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	135262	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.188	152	69758	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	38768	4.617	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	92.400%
7) Chloroethane-d5	1.532	69	34249	4.618	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	92.400%
11) 1,1-Dichloroethene-d2	2.056	65	18216	4.661	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	93.200%
20) 2-Butanone-d5	3.802	46	122078	55.607	ug/L	-0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	111.220%
24) Chloroform-d	4.253	84	82769	4.910	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.200%
26) 1,2-Dichloroethane-d4	4.947	65	46607	5.432	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	108.600%
32) Benzene-d6	4.963	84	150064	4.932	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.600%
36) 1,2-Dichloropropane-d6	5.992	67	46066	5.070	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	101.400%
41) Toluene-d8	7.246	98	129878	4.760	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.200%
43) trans-1,3-Dichloroprop...	7.561	79	15830	4.677	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	93.600%
46) 2-Hexanone-d5	8.027	63	111683	58.355	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	116.700%
56) 1,1,2,2-Tetrachloroeth...	10.156	84	37579	5.226	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	104.600%
66) 1,2-Dichlorobenzene-d4	11.561	152	54860	5.176	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	103.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV022624\
Data File : VV034415.D
Acq On : 26 Feb 2024 18:19
Operator : SY/MD
Sample : P1535-01
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0AB4

Quant Time: Feb 27 05:07:36 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR022624WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Feb 27 05:05:59 2024
Response via : Initial Calibration

