

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV022624\
 Data File : VV034417.D
 Acq On : 26 Feb 2024 19:07
 Operator : SY/MD
 Sample : P1535-04
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AE7

Quant Time: Feb 27 05:08:14 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR022624WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Feb 27 05:05:59 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.538	114	134968	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	127867	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.188	152	64035	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	37789	4.823	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	96.400%	
7) Chloroethane-d5	1.532	69	33083	4.781	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	95.600%	
11) 1,1-Dichloroethene-d2	2.059	65	17359	4.760	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	95.200%	
20) 2-Butanone-d5	3.802	46	111062	54.218	ug/L	-0.02
Spiked Amount	50.000	Range 40 - 130	Recovery	=	108.440%	
24) Chloroform-d	4.252	84	78914	5.017	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	100.400%	
26) 1,2-Dichloroethane-d4	4.947	65	42989	5.370	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	107.400%	
32) Benzene-d6	4.963	84	142414	4.951	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	99.000%	
36) 1,2-Dichloropropane-d6	5.992	67	44993	5.239	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	104.800%	
41) Toluene-d8	7.246	98	125454	4.863	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	97.200%	
43) trans-1,3-Dichloroprop...	7.561	79	15371	4.804	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	96.000%	
46) 2-Hexanone-d5	8.024	63	103485	57.199	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	114.400%	
56) 1,1,2,2-Tetrachloroeth...	10.156	84	36324	5.343	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	106.800%	
66) 1,2-Dichlorobenzene-d4	11.561	152	52315	5.377	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	107.600%	
Target Compounds						
					Qvalue	
10) 1,1,2-Trichloro-1,2,2-...	2.066	101	3582	0.408	ug/L #	86
12) 1,1-Dichloroethene	2.069	96	23178	2.742	ug/L #	82
19) 1,1-Dichloroethane	3.120	63	7043	0.426	ug/L	98
22) cis-1,2-Dichloroethene	3.818	96	21100	2.232	ug/L	95
25) Chloroform	4.275	83	4743	0.276	ug/L	99
31) Carbon tetrachloride	4.738	117	4056	0.292	ug/L	95
34) Trichloroethene	5.834	95	99399	10.756	ug/L	99
47) Tetrachloroethene	7.905	164	55127	7.604	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV022624\
Data File : VV034417.D
Acq On : 26 Feb 2024 19:07
Operator : SY/MD
Sample : P1535-04
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0AE7

Quant Time: Feb 27 05:08:14 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR022624WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Feb 27 05:05:59 2024
Response via : Initial Calibration

