

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW022624\
 Data File : VW034422.D
 Acq On : 26 Feb 2024 21:07
 Operator : SY/MD
 Sample : P1535-10
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AF7

Quant Time: Feb 27 05:09:34 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR022624WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Feb 27 05:05:59 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.538	114	139312	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	133047	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.188	152	67258	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	37521	4.640	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	92.800%
7) Chloroethane-d5	1.532	69	33355	4.670	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	93.400%
11) 1,1-Dichloroethene-d2	2.060	65	17124	4.549	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	91.000%
20) 2-Butanone-d5	3.802	46	114580	54.191	ug/L	-0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	108.380%
24) Chloroform-d	4.252	84	77833	4.794	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.800%
26) 1,2-Dichloroethane-d4	4.947	65	43577	5.274	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	105.400%
32) Benzene-d6	4.963	84	143663	4.800	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.000%
36) 1,2-Dichloropropane-d6	5.992	67	45197	5.057	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	101.200%
41) Toluene-d8	7.246	98	125536	4.677	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.600%
43) trans-1,3-Dichloroprop...	7.561	79	14393	4.323	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	86.400%
46) 2-Hexanone-d5	8.027	63	103279	54.862	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	109.720%
56) 1,1,2,2-Tetrachloroeth...	10.156	84	36558	5.168	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	103.400%
66) 1,2-Dichlorobenzene-d4	11.561	152	53267	5.213	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	104.200%
Target Compounds						
22) cis-1,2-Dichloroethene	3.818	96	5797	0.594	ug/L	95
25) Chloroform	4.278	83	11687	0.659	ug/L	98
31) Carbon tetrachloride	4.738	117	12911	0.894	ug/L	99
34) Trichloroethene	5.841	95	16932	1.761	ug/L	94
47) Tetrachloroethene	7.908	164	28247	3.744	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV022624\
Data File : VV034422.D
Acq On : 26 Feb 2024 21:07
Operator : SY/MD
Sample : P1535-10
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
COAF7

Quant Time: Feb 27 05:09:34 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR022624WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Feb 27 05:05:59 2024
Response via : Initial Calibration

