

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV022819\
 Data File : VV009427.D
 Acq On : 28 Feb 2019 16:56
 Operator : SY/MD
 Sample : K1599-06
 Misc : 25ML/MSVOA V/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 CWJX7

Quant Time: Mar 01 04:20:22 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR022819WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Mar 01 03:49:54 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	95648	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	92123	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	44026	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	25679	5.14	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	102.80%
7) Chloroethane-d5	1.59	69	21943	4.86	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	97.20%
11) 1,1-Dichloroethene-d2	2.13	63	48777	3.86	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	77.20%
20) 2-Butanone-d5	3.95	46	54318	57.34	ug/L	-0.03
Spiked Amount	50.000	Range	40 - 130	Recovery	=	114.68%
24) Chloroform-d	4.41	84	52273	5.34	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	106.80%
26) 1,2-Dichloroethane-d4	5.09	65	26984	5.59	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	111.80%
32) Benzene-d6	5.10	84	106621	5.44	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	108.80%
36) 1,2-Dichloropropane-d6	6.12	67	31330	5.26	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	105.20%
41) Toluene-d8	7.36	98	107833	5.39	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	107.80%
43) trans-1,3-Dichloropropene-	7.67	79	13359	5.07	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	101.40%
46) 2-Hexanone-d5	8.14	63	50191	58.68	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	117.36%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	23863	5.15	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	103.00%
64) 1,2-Dichlorobenzene-d4	11.68	152	37648	5.32	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	106.40%

Target Compounds

					Ovalue
13) Acetone	2.20	43	2418	2.687 ug/L	95
22) cis-1,2-Dichloroethene	3.96	96	79056	12.472 ug/L	99
34) Trichloroethene	5.96	95	269463	39.852 ug/L	98
42) Toluene	7.43	91	21976	0.806 ug/L	98
47) Tetrachloroethene	8.02	164	4788	0.902 ug/L	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV022819\
Data File : VV009427.D
Acq On : 28 Feb 2019 16:56
Operator : SY/MD
Sample : K1599-06
Misc : 25ML/MSVOA V/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
CWJX7

Quant Time: Mar 01 04:20:22 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR022819WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Mar 01 03:49:54 2019
Response via : Initial Calibration

