

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV022819\
 Data File : VV009431.D
 Acq On : 28 Feb 2019 18:39
 Operator : SY/MD
 Sample : K1599-10
 Misc : 25ML/MSVOA V/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 CWJZ0

Quant Time: Mar 01 04:41:52 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR022819WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Mar 01 03:49:54 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	95682	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	90544	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	42107	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	25839	5.17	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	103.40%
7) Chloroethane-d5	1.58	69	23596	5.23	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	104.60%
11) 1,1-Dichloroethene-d2	2.13	63	50771	4.02	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	80.40%
20) 2-Butanone-d5	3.96	46	53313	56.25	ug/L	-0.03
Spiked Amount	50.000	Range	40 - 130	Recovery	=	112.50%
24) Chloroform-d	4.40	84	51899	5.30	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	106.00%
26) 1,2-Dichloroethane-d4	5.09	65	27575	5.71	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	114.20%
32) Benzene-d6	5.10	84	110530	5.74	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	114.80%
36) 1,2-Dichloropropane-d6	6.12	67	32438	5.54	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	110.80%
41) Toluene-d8	7.36	98	111134	5.65	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	113.00%
43) trans-1,3-Dichloropropene-	7.67	79	13236	5.11	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	102.20%
46) 2-Hexanone-d5	8.14	63	48455	57.64	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	115.28%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	25086	5.51	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	110.20%
64) 1,2-Dichlorobenzene-d4	11.68	152	37656	5.56	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	111.20%

Target Compounds

					Ovalue
5) Vinyl chloride	1.32	62	9496	1.392 ug/L	# 1
13) Acetone	2.20	43	3597	3.995 ug/L	87
18) trans-1,2-Dichloroethene	2.79	96	130724	19.252 ug/L	92
22) cis-1,2-Dichloroethene	3.96	96	1119468	176.542 ug/L	91
34) Trichloroethene	5.96	95	64847	9.758 ug/L	98
42) Toluene	7.43	91	36814	1.373 ug/L	98
51) Chlorobenzene	8.93	112	3431	0.196 ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV022819\
Data File : VV009431.D
Acq On : 28 Feb 2019 18:39
Operator : SY/MD
Sample : K1599-10
Misc : 25ML/MSVOA V/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
CWJZ0

Quant Time: Mar 01 04:41:52 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR022819WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Mar 01 03:49:54 2019
Response via : Initial Calibration

