

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV022819\
 Data File : VV009432.D
 Acq On : 28 Feb 2019 19:05
 Operator : SY/MD
 Sample : K1599-11
 Misc : 25ML/MSVOA V/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 CWJZ1

Quant Time: Mar 01 04:44:43 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR022819WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Mar 01 03:49:54 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	87674	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	82180	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	38846	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	20630	4.51	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	90.20%
7) Chloroethane-d5	1.58	69	22716	5.49	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	109.80%
11) 1,1-Dichloroethene-d2	2.13	63	46270	3.99	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	79.80%
20) 2-Butanone-d5	3.96	46	54956	63.28	ug/L	-0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	126.56%
24) Chloroform-d	4.40	84	51775	5.77	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	115.40%
26) 1,2-Dichloroethane-d4	5.09	65	26347	5.96	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	119.20%
32) Benzene-d6	5.10	84	106916	6.12	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	122.40%
36) 1,2-Dichloropropane-d6	6.12	67	31839	5.99	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	119.80%
41) Toluene-d8	7.36	98	104152	5.83	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	116.60%
43) trans-1,3-Dichloropropene-	7.67	79	8941	3.80	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	76.00%
46) 2-Hexanone-d5	8.14	63	48698	63.83	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	127.66%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	23852	5.77	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	115.40%
64) 1,2-Dichlorobenzene-d4	11.68	152	38870	6.22	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	124.40%#

Target Compounds

					Ovalue
13) Acetone	2.21	43	3014	3.653 ug/L	95
22) cis-1,2-Dichloroethene	3.97	96	6426	1.106 ug/L	92
34) Trichloroethene	5.96	95	3166	0.525 ug/L	89
42) Toluene	7.43	91	9647	0.397 ug/L	91
47) Tetrachloroethene	8.02	164	2068	0.437 ug/L #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV022819\
Data File : VV009432.D
Acq On : 28 Feb 2019 19:05
Operator : SY/MD
Sample : K1599-11
Misc : 25ML/MSVOA V/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
CWJZ1

Quant Time: Mar 01 04:44:43 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR022819WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Mar 01 03:49:54 2019
Response via : Initial Calibration

