

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW022823\
 Data File : VW030094.D
 Acq On : 28 Feb 2023 13:05
 Operator : SY/MD
 Sample : 01729-06
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0BZ5

Quant Time: Mar 01 00:06:13 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR022323WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Mar 01 00:04:57 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.545	114	208495	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.792	117	196485	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.194	152	103051	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.285	65	56632	4.198	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	84.000%
7) Chloroethane-d5	1.539	69	55529	4.385	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	87.600%
11) 1,1-Dichloroethene-d2	2.066	63	89424	3.224	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	64.400%
20) 2-Butanone-d5	3.822	46	112382	45.183	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	90.360%
24) Chloroform-d	4.262	84	117796	4.274	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	85.400%
26) 1,2-Dichloroethane-d4	4.953	65	52514	4.668	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.400%
32) Benzene-d6	4.973	84	238057	4.503	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	90.000%
36) 1,2-Dichloropropane-d6	5.998	67	73896	4.646	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	93.000%
41) Toluene-d8	7.252	98	213597	4.262	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	85.200%
43) trans-1,3-Dichloroprop...	7.561	79	24521	4.506	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	90.200%
46) 2-Hexanone-d5	8.034	63	104846	46.571	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	93.140%
56) 1,1,2,2-Tetrachloroeth...	10.162	84	51893	4.586	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	91.800%
66) 1,2-Dichlorobenzene-d4	11.570	152	84632	4.854	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	97.000%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV022823\
Data File : VV030094.D
Acq On : 28 Feb 2023 13:05
Operator : SY/MD
Sample : 01729-06
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
COBZ5

Quant Time: Mar 01 00:06:13 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR022323WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Mar 01 00:04:57 2023
Response via : Initial Calibration

