

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\Data\VW022823\  
 Data File : VW030096.D  
 Acq On : 28 Feb 2023 13:52  
 Operator : SY/MD  
 Sample : 01729-02  
 Misc : 25.0mL/MSVOA\_V/WATER  
 ALS Vial : 8 Sample Multiplier: 1

Instrument :  
 MSVOA\_V  
 ClientSampleId :  
 C0BY6

Quant Time: Mar 01 00:06:41 2023  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\SFAMVTR022323WMA.M  
 Quant Title : TRACE VOA SFAM1.0  
 QLast Update : Wed Mar 01 00:04:57 2023  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon  | Response | Conc     | Units | Dev(Min)  |
|-------------------------------|--------|-------|----------|----------|-------|-----------|
| -----                         |        |       |          |          |       |           |
| Internal Standards            |        |       |          |          |       |           |
| 1) 1,4-Difluorobenzene        | 5.545  | 114   | 181077   | 5.000    | ug/L  | 0.00      |
| 28) Chlorobenzene-d5          | 8.792  | 117   | 172719   | 5.000    | ug/L  | 0.00      |
| 58) 1,4-Dichlorobenzene-d4    | 11.194 | 152   | 90578    | 5.000    | ug/L  | 0.00      |
| System Monitoring Compounds   |        |       |          |          |       |           |
| 4) Vinyl Chloride-d3          | 1.281  | 65    | 54734    | 4.671    | ug/L  | 0.00      |
| Spiked Amount                 | 5.000  | Range | 40 - 130 | Recovery | =     | 93.400%   |
| 7) Chloroethane-d5            | 1.539  | 69    | 54336    | 4.940    | ug/L  | 0.00      |
| Spiked Amount                 | 5.000  | Range | 65 - 130 | Recovery | =     | 98.800%   |
| 11) 1,1-Dichloroethene-d2     | 2.066  | 63    | 85411    | 3.546    | ug/L  | 0.00      |
| Spiked Amount                 | 5.000  | Range | 60 - 125 | Recovery | =     | 71.000%   |
| 20) 2-Butanone-d5             | 3.822  | 46    | 109916   | 50.883   | ug/L  | 0.00      |
| Spiked Amount                 | 50.000 | Range | 40 - 130 | Recovery | =     | 101.760%  |
| 24) Chloroform-d              | 4.262  | 84    | 117761   | 4.920    | ug/L  | 0.00      |
| Spiked Amount                 | 5.000  | Range | 70 - 125 | Recovery | =     | 98.400%   |
| 26) 1,2-Dichloroethane-d4     | 4.957  | 65    | 51634    | 5.285    | ug/L  | 0.00      |
| Spiked Amount                 | 5.000  | Range | 70 - 130 | Recovery | =     | 105.600%  |
| 32) Benzene-d6                | 4.973  | 84    | 230605   | 4.962    | ug/L  | 0.00      |
| Spiked Amount                 | 5.000  | Range | 70 - 125 | Recovery | =     | 99.200%   |
| 36) 1,2-Dichloropropane-d6    | 5.998  | 67    | 71287    | 5.099    | ug/L  | 0.00      |
| Spiked Amount                 | 5.000  | Range | 60 - 140 | Recovery | =     | 102.000%  |
| 41) Toluene-d8                | 7.252  | 98    | 206510   | 4.688    | ug/L  | 0.00      |
| Spiked Amount                 | 5.000  | Range | 70 - 130 | Recovery | =     | 93.800%   |
| 43) trans-1,3-Dichloroprop... | 7.561  | 79    | 24009    | 5.019    | ug/L  | 0.00      |
| Spiked Amount                 | 5.000  | Range | 55 - 130 | Recovery | =     | 100.400%  |
| 46) 2-Hexanone-d5             | 8.034  | 63    | 106185   | 53.656   | ug/L  | 0.00      |
| Spiked Amount                 | 50.000 | Range | 45 - 130 | Recovery | =     | 107.320%  |
| 56) 1,1,2,2-Tetrachloroeth... | 10.162 | 84    | 51670    | 5.195    | ug/L  | 0.00      |
| Spiked Amount                 | 5.000  | Range | 65 - 120 | Recovery | =     | 103.800%  |
| 66) 1,2-Dichlorobenzene-d4    | 11.570 | 152   | 82370    | 5.375    | ug/L  | 0.00      |
| Spiked Amount                 | 5.000  | Range | 80 - 120 | Recovery | =     | 107.400%  |
| Target Compounds              |        |       |          |          |       |           |
| 25) Chloroform                | 4.288  | 83    | 18628    | 0.832    | ug/L  | Qvalue 93 |
| -----                         |        |       |          |          |       |           |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\Data\VV022823\  
Data File : VV030096.D  
Acq On : 28 Feb 2023 13:52  
Operator : SY/MD  
Sample : 01729-02  
Misc : 25.0mL/MSVOA\_V/WATER  
ALS Vial : 8 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
C0BY6

Quant Time: Mar 01 00:06:41 2023  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\SFAMVTR022323WMA.M  
Quant Title : TRACE VOA SFAM1.0  
QLast Update : Wed Mar 01 00:04:57 2023  
Response via : Initial Calibration

