




Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV022824\
Data File : VV034463.D
Acq On : 28 Feb 2024 11:53
Operator : SY/MD
Sample : P1566-06DL 5X
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
JWAC0DL

Manual IntegrationsAPPROVED

Quant Time: Feb 29 00:36:44 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR022624WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Feb 29 00:34:49 2024
Response via : Initial Calibration

Reviewed By :Mahesh Dadoda 02/29/2024
Supervised By :Semsettin Yesilyurt 02/29/2024



TIC: VV034463.D\data.ms

(13) Acetone (T)

2.163min (+ 0.019) 47.08 ug/L

response 102743

Ion	Exp%	Act%
43.05	100.00	100.00
58.05	28.10	28.75
0.00	0.00	0.00
0.00	0.00	0.00




Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV022824\
Data File : VV034463.D
Acq On : 28 Feb 2024 11:53
Operator : SY/MD
Sample : P1566-06DL 5X
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
JWAC0DL

Manual IntegrationsAPPROVED

Quant Time: Feb 29 00:36:44 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR022624WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Feb 29 00:34:49 2024
Response via : Initial Calibration

Reviewed By :Mahesh Dadoda 02/29/2024
Supervised By :Semsettin Yesilyurt 02/29/2024



TIC: VV034463.D\data.ms

(13) Acetone (T)

2.163min (+ 0.019) 51.44 ug/L m

response 112263

Ion	Exp%	Act%
43.05	100.00	100.00
58.05	28.10	26.31
0.00	0.00	0.00
0.00	0.00	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV022824\
 Data File : VV034463.D
 Acq On : 28 Feb 2024 11:53
 Operator : SY/MD
 Sample : P1566-06DL 5X
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 JWAC0DL

Manual IntegrationsAPPROVED

Quant Time: Feb 29 00:36:44 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR022624WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Feb 29 00:34:49 2024
 Response via : Initial Calibration

Reviewed By :Mahesh Dadoda 02/29/2024
 Supervised By :Semsettin Yesilyurt 02/29/2024

Compound	R.T.	Q	Ion	Response	Conc	Units	Dev(Min)

Internal Standards							
1) 1,4-Difluorobenzene	5.535	114		141378	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117		135171	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.188	152		67737	5.000	ug/L	0.00
System Monitoring Compounds							
4) Vinyl Chloride-d3	1.278	65		31688	3.861	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	77.200%	
7) Chloroethane-d5	1.532	69		29776	4.108	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	82.200%	
11) 1,1-Dichloroethene-d2	2.060	65		14938	3.910	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	78.200%	
20) 2-Butanone-d5	3.828	46		114427	53.328	ug/L	0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	106.660%	
24) Chloroform-d	4.253	84		76050	4.615	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.400%	
26) 1,2-Dichloroethane-d4	4.947	65		41391	4.936	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.800%	
32) Benzene-d6	4.963	84		134837	4.434	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.600%	
36) 1,2-Dichloropropane-d6	5.992	67		43296	4.769	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	95.400%	
41) Toluene-d8	7.246	98		115214	4.225	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	84.600%	
43) trans-1,3-Dichloroprop...	7.561	79		15405	4.554	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	91.000%	
46) 2-Hexanone-d5	8.031	63		104072	54.415	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	108.820%	
56) 1,1,2,2-Tetrachloroeth...	10.156	84		35685	4.966	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	99.400%	
66) 1,2-Dichlorobenzene-d4	11.561	152		49809	4.840	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	96.800%	
Target Compounds							
13) Acetone	2.163	43		112263m	51.440	ug/L	
47) Tetrachloroethene	7.912	164		1721	0.225	ug/L	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV022824\
Data File : VV034463.D
Acq On : 28 Feb 2024 11:53
Operator : SY/MD
Sample : P1566-06DL 5X
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
JWAC0DL

Manual IntegrationsAPPROVED

Quant Time: Feb 29 00:36:44 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR022624WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Feb 29 00:34:49 2024
Response via : Initial Calibration

Reviewed By :Mahesh Dadoda 02/29/2024
Supervised By :Semsettin Yesilyurt 02/29/2024

