

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV030119\
 Data File : VV009448.D
 Acq On : 01 Mar 2019 11:58
 Operator : SY/MD
 Sample : K1599-11RE
 Misc : 25ML/MSVOA V/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 CWJZ1RE

Quant Time: Mar 02 00:28:28 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR022819WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Mar 02 00:06:51 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	91176	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	86118	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	39685	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	23901	5.02	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	100.40%
7) Chloroethane-d5	1.58	69	22170	5.15	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	103.00%
11) 1,1-Dichloroethene-d2	2.13	63	46331	3.85	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	77.00%
20) 2-Butanone-d5	3.97	46	52293	57.90	ug/L	0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	115.80%
24) Chloroform-d	4.41	84	49329	5.29	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	105.80%
26) 1,2-Dichloroethane-d4	5.09	65	25222	5.48	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	109.60%
32) Benzene-d6	5.10	84	101402	5.54	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	110.80%
36) 1,2-Dichloropropane-d6	6.12	67	30874	5.55	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	111.00%
41) Toluene-d8	7.36	98	99959	5.34	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	106.80%
43) trans-1,3-Dichloropropene-	7.67	79	12612	5.12	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	102.40%
46) 2-Hexanone-d5	8.14	63	45963	57.49	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	114.98%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	23010	5.31	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	106.20%
64) 1,2-Dichlorobenzene-d4	11.68	152	36281	5.68	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	113.60%

Target Compounds

					Ovalue
13) Acetone	2.22	43	4133	4.817 ug/L	90
22) cis-1,2-Dichloroethene	3.96	96	5185	0.858 ug/L	87
34) Trichloroethene	5.96	95	2930	0.464 ug/L	92
42) Toluene	7.43	91	9722	0.381 ug/L	89
47) Tetrachloroethene	8.02	164	2177	0.439 ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV030119\
Data File : VV009448.D
Acq On : 01 Mar 2019 11:58
Operator : SY/MD
Sample : K1599-11RE
Misc : 25ML/MSVOA V/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
CWJZ1RE

Quant Time: Mar 02 00:28:28 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR022819WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Mar 02 00:06:51 2019
Response via : Initial Calibration

