

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV030119\
 Data File : VV009451.D
 Acq On : 01 Mar 2019 13:15
 Operator : SY/MD
 Sample : K1599-08DL 5X
 Misc : 25ML/MSVOA V/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 CWJY6DL

Quant Time: Mar 02 00:49:37 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR022819WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Mar 02 00:06:51 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	89174	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	83088	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	39294	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	21298	4.58	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	91.60%
7) Chloroethane-d5	1.57	69	20742	4.93	ug/L	-0.02
Spiked Amount	5.000	Range	65 - 130	Recovery	=	98.60%
11) 1,1-Dichloroethene-d2	2.13	63	44524	3.78	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	75.60%
20) 2-Butanone-d5	3.96	46	54625	61.85	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	123.70%
24) Chloroform-d	4.40	84	47476	5.21	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	104.20%
26) 1,2-Dichloroethane-d4	5.09	65	25392	5.64	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	112.80%
32) Benzene-d6	5.10	84	97233	5.50	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	110.00%
36) 1,2-Dichloropropane-d6	6.12	67	30752	5.73	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	114.60%
41) Toluene-d8	7.36	98	97281	5.39	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	107.80%
43) trans-1,3-Dichloropropene-	7.67	79	11994	5.04	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	100.80%
46) 2-Hexanone-d5	8.14	63	51819	67.18	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	134.36%#
57) 1,1,2,2-Tetrachloroethane-	10.26	84	22219	5.32	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	106.40%
64) 1,2-Dichlorobenzene-d4	11.68	152	34529	5.46	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	109.20%

Target Compounds					Ovalue
18) trans-1,2-Dichloroethene	2.78	96	3457	0.546 ug/L	89
22) cis-1,2-Dichloroethene	3.96	96	77069	13.041 ug/L	98
34) Trichloroethene	5.96	95	63395	10.395 ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV030119\
Data File : VV009451.D
Acq On : 01 Mar 2019 13:15
Operator : SY/MD
Sample : K1599-08DL 5X
Misc : 25ML/MSVOA V/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
CWJY6DL

Quant Time: Mar 02 00:49:37 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR022819WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Mar 02 00:06:51 2019
Response via : Initial Calibration

