

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV030119\
 Data File : VV009453.D
 Acq On : 01 Mar 2019 14:06
 Operator : SY/MD
 Sample : K1599-18DL 5X
 Misc : 25ML/MSVOA V/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 CWK05DL

Quant Time: Mar 02 00:58:09 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR022819WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Mar 02 00:06:51 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	87357	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	81520	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	38794	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	22796	5.00	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	100.00%
7) Chloroethane-d5	1.58	69	20639	5.01	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	100.20%
11) 1,1-Dichloroethene-d2	2.13	63	42812	3.71	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	74.20%
20) 2-Butanone-d5	3.96	46	50308	58.14	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	116.28%
24) Chloroform-d	4.40	84	46407	5.19	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	103.80%
26) 1,2-Dichloroethane-d4	5.08	65	24332	5.52	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	110.40%
32) Benzene-d6	5.10	84	96712	5.58	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	111.60%
36) 1,2-Dichloropropane-d6	6.12	67	29409	5.58	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	111.60%
41) Toluene-d8	7.36	98	97269	5.49	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	109.80%
43) trans-1,3-Dichloropropene-	7.67	79	11392	4.88	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	97.60%
46) 2-Hexanone-d5	8.14	63	46027	60.82	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	121.64%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	23414	5.71	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	114.20%
64) 1,2-Dichlorobenzene-d4	11.68	152	35061	5.62	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	112.40%

Target Compounds

					Ovalue
22) cis-1,2-Dichloroethene	3.96	96	15055	2.600 ug/L	100
34) Trichloroethene	5.96	95	48437	8.095 ug/L	97
42) Toluene	7.43	91	9271	0.384 ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV030119\
Data File : VV009453.D
Acq On : 01 Mar 2019 14:06
Operator : SY/MD
Sample : K1599-18DL 5X
Misc : 25ML/MSVOA V/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
CWK05DL

Quant Time: Mar 02 00:58:09 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR022819WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Mar 02 00:06:51 2019
Response via : Initial Calibration

