

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV030119\
 Data File : VV009456.D
 Acq On : 01 Mar 2019 15:24
 Operator : SY/MD
 Sample : K1588-21
 Misc : 25ML/MSVOA V/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK01

Quant Time: Mar 02 01:13:34 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR022819WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Mar 02 00:06:51 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	91056	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	85462	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	40461	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	22215	4.68	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	93.60%
7) Chloroethane-d5	1.58	69	20634	4.80	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	96.00%
11) 1,1-Dichloroethene-d2	2.13	63	44342	3.69	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	73.80%
20) 2-Butanone-d5	3.99	46	49393	54.77	ug/L	0.03
Spiked Amount	50.000	Range	40 - 130	Recovery	=	109.54%
24) Chloroform-d	4.40	84	43578	4.68	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.60%
26) 1,2-Dichloroethane-d4	5.09	65	22812	4.97	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	99.40%
32) Benzene-d6	5.10	84	89970	4.95	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.00%
36) 1,2-Dichloropropane-d6	6.12	67	27776	5.03	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	100.60%
41) Toluene-d8	7.36	98	89255	4.81	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.20%
43) trans-1,3-Dichloropropene-	7.67	79	10468	4.28	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	85.60%
46) 2-Hexanone-d5	8.14	63	42168	53.15	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	106.30%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	20930	4.87	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	97.40%
64) 1,2-Dichlorobenzene-d4	11.68	152	32165	4.94	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	98.80%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV030119\
Data File : VV009456.D
Acq On : 01 Mar 2019 15:24
Operator : SY/MD
Sample : K1588-21
Misc : 25ML/MSVOA V/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK01

Quant Time: Mar 02 01:13:34 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR022819WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Mar 02 00:06:51 2019
Response via : Initial Calibration

