

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV030123\
 Data File : VV030104.D
 Acq On : 01 Mar 2023 12:51
 Operator : SY/MD
 Sample : 01729-09
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0BY1

Quant Time: Mar 02 00:28:30 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR022323WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Mar 02 00:27:39 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.545	114	158632	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.792	117	152660	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.194	152	78126	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.282	65	47947	4.671	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	93.400%	
7) Chloroethane-d5	1.539	69	50144	5.204	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	104.000%	
11) 1,1-Dichloroethene-d2	2.066	63	76163	3.609	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	72.200%	
20) 2-Butanone-d5	3.825	46	101536	53.654	ug/L	-0.03
Spiked Amount	50.000	Range 40 - 130	Recovery	=	107.300%	
24) Chloroform-d	4.262	84	108464	5.173	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	103.400%	
26) 1,2-Dichloroethane-d4	4.953	65	46712	5.457	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	109.200%	
32) Benzene-d6	4.969	84	201658	4.909	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	98.200%	
36) 1,2-Dichloropropane-d6	5.995	67	63419	5.132	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	102.600%	
41) Toluene-d8	7.252	98	173180	4.448	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	89.000%	
43) trans-1,3-Dichloroprop...	7.561	79	20973	4.961	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	99.200%	
46) 2-Hexanone-d5	8.034	63	91478	52.298	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	104.600%	
56) 1,1,2,2-Tetrachloroeth...	10.162	84	46669	5.308	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	106.200%	
66) 1,2-Dichlorobenzene-d4	11.570	152	67423	5.101	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	102.000%	
Target Compounds						
16) Methylene chloride	2.455	84	3581	0.275	ug/L	97
25) Chloroform	4.288	83	103577	5.278	ug/L	96
38) Bromodichloromethane	6.445	83	16269	1.201	ug/L	96
49) Dibromochloromethane	8.191	129	1337	0.151	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV030123\
Data File : VV030104.D
Acq On : 01 Mar 2023 12:51
Operator : SY/MD
Sample : 01729-09
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
COBY1

Quant Time: Mar 02 00:28:30 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR022323WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Mar 02 00:27:39 2023
Response via : Initial Calibration

