

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW030123\
 Data File : VW030109.D
 Acq On : 01 Mar 2023 14:50
 Operator : SY/MD
 Sample : 01729-08
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK001

Quant Time: Mar 02 00:29:50 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR022323WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Mar 02 00:27:39 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.545	114	158583	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.792	117	152852	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.194	152	76644	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.281	65	45336	4.418	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	88.400%	
7) Chloroethane-d5	1.539	69	46450	4.822	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	96.400%	
11) 1,1-Dichloroethene-d2	2.066	63	74082	3.512	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	70.200%	
20) 2-Butanone-d5	3.831	46	95533	50.497	ug/L	-0.02
Spiked Amount	50.000	Range 40 - 130	Recovery	=	101.000%	
24) Chloroform-d	4.262	84	101152	4.826	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	96.600%	
26) 1,2-Dichloroethane-d4	4.957	65	44856	5.242	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	104.800%	
32) Benzene-d6	4.973	84	189215	4.600	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	92.000%	
36) 1,2-Dichloropropane-d6	5.998	67	60806	4.914	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	98.200%	
41) Toluene-d8	7.252	98	158365	4.062	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	81.200%	
43) trans-1,3-Dichloroprop...	7.564	79	18876	4.459	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	89.200%	
46) 2-Hexanone-d5	8.037	63	87977	50.233	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	100.460%	
56) 1,1,2,2-Tetrachloroeth...	10.162	84	44414	5.045	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	101.000%	
66) 1,2-Dichlorobenzene-d4	11.570	152	63287	4.880	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	97.600%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV030123\
Data File : VV030109.D
Acq On : 01 Mar 2023 14:50
Operator : SY/MD
Sample : 01729-08
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK001

Quant Time: Mar 02 00:29:50 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR022323WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Mar 02 00:27:39 2023
Response via : Initial Calibration

