

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW030123\
 Data File : VW030107.D
 Acq On : 01 Mar 2023 14:04
 Operator : SY/MD
 Sample : 01729-12
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0BZ6

Quant Time: Mar 02 00:29:18 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR022323WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Mar 02 00:27:39 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.545	114	159838	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.793	117	153912	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.194	152	77831	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.282	65	50500	4.883	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	97.600%
7) Chloroethane-d5	1.539	69	50909	5.244	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	104.800%
11) 1,1-Dichloroethene-d2	2.066	63	80178	3.771	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	75.400%
20) 2-Butanone-d5	3.818	46	104662	54.888	ug/L	-0.03
Spiked Amount	50.000	Range	40 - 130	Recovery	=	109.780%
24) Chloroform-d	4.262	84	110114	5.212	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	104.200%
26) 1,2-Dichloroethane-d4	4.953	65	49541	5.744	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	114.800%
32) Benzene-d6	4.973	84	208948	5.045	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	101.000%
36) 1,2-Dichloropropane-d6	5.998	67	67651	5.430	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	108.600%
41) Toluene-d8	7.252	98	178336	4.543	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.800%
43) trans-1,3-Dichloroprop...	7.561	79	20359	4.776	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	95.600%
46) 2-Hexanone-d5	8.034	63	97599	55.344	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	110.680%
56) 1,1,2,2-Tetrachloroeth...	10.162	84	49306	5.563	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	111.200%
66) 1,2-Dichlorobenzene-d4	11.571	152	70445	5.350	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	107.000%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV030123\
Data File : VV030107.D
Acq On : 01 Mar 2023 14:04
Operator : SY/MD
Sample : 01729-12
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
COBZ6

Quant Time: Mar 02 00:29:18 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR022323WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Mar 02 00:27:39 2023
Response via : Initial Calibration

