

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW030123\
 Data File : VW030108.D
 Acq On : 01 Mar 2023 14:26
 Operator : SY/MD
 Sample : 01729-13
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0BZ8

Quant Time: Mar 02 00:29:33 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR022323WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Mar 02 00:27:39 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.545	114	153620	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.796	117	149069	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.194	152	75418	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.282	65	46781	4.706	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	94.200%
7) Chloroethane-d5	1.539	69	48525	5.201	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	104.000%
11) 1,1-Dichloroethene-d2	2.066	63	77083	3.772	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	75.400%
20) 2-Butanone-d5	3.835	46	101103	55.168	ug/L	-0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	110.340%
24) Chloroform-d	4.262	84	106920	5.266	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	105.400%
26) 1,2-Dichloroethane-d4	4.957	65	46665	5.630	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	112.600%
32) Benzene-d6	4.976	84	199343	4.970	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.400%
36) 1,2-Dichloropropane-d6	6.002	67	62836	5.207	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	104.200%
41) Toluene-d8	7.256	98	170629	4.488	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	89.800%
43) trans-1,3-Dichloroprop...	7.564	79	19856	4.809	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	96.200%
46) 2-Hexanone-d5	8.037	63	93366	54.663	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	109.320%
56) 1,1,2,2-Tetrachloroeth...	10.162	84	48514	5.651	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	113.000%
66) 1,2-Dichlorobenzene-d4	11.571	152	68447	5.364	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	107.200%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV030123\
Data File : VV030108.D
Acq On : 01 Mar 2023 14:26
Operator : SY/MD
Sample : 01729-13
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
COBZ8

Quant Time: Mar 02 00:29:33 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR022323WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Mar 02 00:27:39 2023
Response via : Initial Calibration

