

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV030221\
 Data File : VV020432.D
 Acq On : 02 Mar 2021 12:20
 Operator : SY/MD
 Sample : VSTD00234
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VSTD002306

Quant Time: Mar 03 00:24:57 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM030221.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Mar 03 00:23:40 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.639	114	12822	0.50	ug/L	0.00
5) Chlorobenzene-d5	8.864	117	12582	0.50	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.261	152	6345	0.50	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.303	65	30626	1.90	ug/L	0.00
4) 1,2-Dichloroethane-d4	5.042	65	18098	1.88	ug/L	0.00
7) 1,2-Dichloropropane-d6	6.069	67	21796	1.90	ug/L	0.00
8) Toluene-d8	7.317	98	66237	2.05	ug/L	0.00
10) 1,1,2,2-Tetrachloroeth...	10.216	84	11380	1.70	ug/L	0.00
Target Compounds						
						Qvalue
3) Vinyl chloride	1.310	62	28151	1.95	ug/L	100
6) Trichloroethene	5.925	95	20371	2.01	ug/L	99
9) 1,2-Dibromoethane	8.353	107	11762	2.00	ug/L #	95
12) 1,2,3-Trichloropropane	10.267	75	10403	1.94	ug/L	100
13) 1,2-Dibromo-3-chloropr...	12.427	75	1337	1.48	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV030221\
Data File : VV020432.D
Acq On : 02 Mar 2021 12:20
Operator : SY/MD
Sample : VSTD00234
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD002306

Quant Time: Mar 03 00:24:57 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM030221.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Mar 03 00:23:40 2021
Response via : Initial Calibration

