

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV030221\
 Data File : VV020429.D
 Acq On : 02 Mar 2021 11:07
 Operator : SY/MD
 Sample : VSTD0.131
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VSTD0.1303

Quant Time: Mar 03 00:24:26 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM030221.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Mar 03 00:23:40 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.639	114	12321	0.50	ug/L	0.00
5) Chlorobenzene-d5	8.863	117	12077	0.50	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.261	152	6088	0.50	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.310	65	1370	0.09	ug/L	0.00
4) 1,2-Dichloroethane-d4	5.042	65	792	0.09	ug/L	0.00
7) 1,2-Dichloropropane-d6	6.068	67	926	0.08	ug/L	0.00
8) Toluene-d8	7.332	98	2799	0.09	ug/L	0.02
10) 1,1,2,2-Tetrachloroeth...	10.216	84	505	0.08	ug/L	0.00
Target Compounds						
						Qvalue
3) Vinyl chloride	1.310	62	1434	0.10	ug/L	92
6) Trichloroethene	5.925	95	984	0.10	ug/L	94
9) 1,2-Dibromoethane	8.368	107	562	0.10	ug/L #	96
12) 1,2,3-Trichloropropane	10.267	75	503	0.10	ug/L	98
13) 1,2-Dibromo-3-chloropr...	12.426	75	69	0.08	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV030221\
Data File : VV020429.D
Acq On : 02 Mar 2021 11:07
Operator : SY/MD
Sample : VSTD0.131
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD0.1303

Quant Time: Mar 03 00:24:26 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM030221.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Mar 03 00:23:40 2021
Response via : Initial Calibration

