

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV030221\
 Data File : VV020431.D
 Acq On : 02 Mar 2021 11:56
 Operator : SY/MD
 Sample : VSTD00133
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VSTD001305

Quant Time: Mar 03 00:24:46 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM030221.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Mar 03 00:23:40 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc Units	Dev(Min)
Internal Standards					
1) 1,4-Difluorobenzene	5.639	114	13026	0.50 ug/L	0.00
5) Chlorobenzene-d5	8.863	117	12727	0.50 ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.261	152	6370	0.50 ug/L	0.00
System Monitoring Compounds					
2) Vinyl Chloride-d3	1.303	65	16020	0.98 ug/L	0.00
4) 1,2-Dichloroethane-d4	5.042	65	9344	0.96 ug/L	0.00
7) 1,2-Dichloropropane-d6	6.068	67	11271	0.97 ug/L	0.00
8) Toluene-d8	7.317	98	33606	1.03 ug/L	0.00
10) 1,1,2,2-Tetrachloroeth...	10.216	84	5905	0.87 ug/L	0.00
Target Compounds					
3) Vinyl chloride	1.310	62	14931	1.02 ug/L	Qvalue 100
6) Trichloroethene	5.925	95	10570	1.03 ug/L	99
9) 1,2-Dibromoethane	8.353	107	6138	1.03 ug/L #	95
12) 1,2,3-Trichloropropane	10.267	75	5476	1.02 ug/L	99
13) 1,2-Dibromo-3-chloropr...	12.426	75	714	0.79 ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV030221\
Data File : VV020431.D
Acq On : 02 Mar 2021 11:56
Operator : SY/MD
Sample : VSTD00133
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD001305

Quant Time: Mar 03 00:24:46 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM030221.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Mar 03 00:23:40 2021
Response via : Initial Calibration

