

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV030221\
 Data File : VV020434.D
 Acq On : 02 Mar 2021 13:39
 Operator : SY/MD
 Sample : VV0302WBL01
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK312

Quant Time: Mar 03 00:35:19 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM030221.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Mar 03 00:31:48 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.639	114	11893	0.50	ug/L	0.00
5) Chlorobenzene-d5	8.863	117	11710	0.50	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.261	152	5750	0.50	ug/L	0.00

System Monitoring Compounds						
2) Vinyl Chloride-d3	1.310	65	7635	0.56	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	=	112.00%
4) 1,2-Dichloroethane-d4	5.042	65	4352	0.55	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	110.00%
7) 1,2-Dichloropropane-d6	6.068	67	5419	0.56	ug/L	0.00
Spiked Amount	0.500	Range	60 - 140	Recovery	=	112.00%
8) Toluene-d8	7.317	98	15184	0.52	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	104.00%
10) 1,1,2,2-Tetrachloroeth...	10.216	84	2783	0.54	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	=	108.00%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV030221\
Data File : VV020434.D
Acq On : 02 Mar 2021 13:39
Operator : SY/MD
Sample : VV0302WBL01
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK312

Quant Time: Mar 03 00:35:19 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM030221.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Mar 03 00:31:48 2021
Response via : Initial Calibration

