

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV030223\
 Data File : VV030113.D
 Acq On : 02 Mar 2023 10:31
 Operator : SY/MD
 Sample : VV0302WBL01
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK306

Manual Integrations
 APPROVED

Reviewed By :Krupa Patel 03/03/2023
 Supervised By :Mahesh Dadoda 03/03/2023

Quant Time: Mar 03 03:30:34 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR022323WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Mar 03 03:28:37 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.545	114	143290	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.793	117	142339	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.194	152	70046	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.282	65	44711	4.822	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	96.400%	
7) Chloroethane-d5	1.536	69	46921	5.391	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	107.800%	
11) 1,1-Dichloroethene-d2	2.063	63	72735	3.816	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	76.400%	
20) 2-Butanone-d5	3.851	46	85840m	50.216	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	100.440%	
24) Chloroform-d	4.259	84	100255	5.293	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	105.800%	
26) 1,2-Dichloroethane-d4	4.953	65	44564	5.764	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	115.200%	
32) Benzene-d6	4.970	84	183644	4.795	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	95.800%	
36) 1,2-Dichloropropane-d6	5.998	67	60134	5.219	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	104.400%	
41) Toluene-d8	7.252	98	154267	4.249	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	85.000%	
43) trans-1,3-Dichloroprop...	7.564	79	17997	4.565	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	91.400%	
46) 2-Hexanone-d5	8.040	63	82294	50.459	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	100.920%	
56) 1,1,2,2-Tetrachloroeth...	10.162	84	45176	5.511	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	110.200%	
66) 1,2-Dichlorobenzene-d4	11.571	152	61439	5.184	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	103.600%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV030223\
Data File : VV030113.D
Acq On : 02 Mar 2023 10:31
Operator : SY/MD
Sample : VV0302WBL01
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK306

Quant Time: Mar 03 03:30:34 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR022323WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Mar 03 03:28:37 2023
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Krupa Patel 03/03/2023
Supervised By :Mahesh Dadoda 03/03/2023

