

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV030321\
 Data File : VV020439.D
 Acq On : 03 Mar 2021 10:09
 Operator : SY/MD
 Sample : VV0303WBL01
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK3213

Quant Time: Mar 04 00:24:06 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM030221.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Mar 04 00:23:40 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc Units	Dev(Min)
Internal Standards					
1) 1,4-Difluorobenzene	5.639	114	12344	0.50 ug/L	0.00
5) Chlorobenzene-d5	8.863	117	11774	0.50 ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.261	152	5801	0.50 ug/L	0.00

System Monitoring Compounds					
2) Vinyl Chloride-d3	1.310	65	7257	0.51 ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	= 102.00%
4) 1,2-Dichloroethane-d4	5.042	65	4197	0.51 ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	= 102.00%
7) 1,2-Dichloropropane-d6	6.068	67	5177	0.53 ug/L	0.00
Spiked Amount	0.500	Range	60 - 140	Recovery	= 106.00%
8) Toluene-d8	7.317	98	14343	0.49 ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	= 98.00%
10) 1,1,2,2-Tetrachloroeth...	10.216	84	2518	0.48 ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	= 96.00%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV030321\
Data File : VV020439.D
Acq On : 03 Mar 2021 10:09
Operator : SY/MD
Sample : VV0303WBL01
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK3213

Quant Time: Mar 04 00:24:06 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM030221.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Mar 04 00:23:40 2021
Response via : Initial Calibration

