

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV030321\
 Data File : VV020438.D
 Acq On : 03 Mar 2021 09:35
 Operator : SY/MD
 Sample : VSTDCCC0.5
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VSTD0.5313

Quant Time: Mar 04 00:17:01 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM030221.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Mar 03 00:31:48 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.639	114	13048	0.50	ug/L	0.00
5) Chlorobenzene-d5	8.863	117	12547	0.50	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.261	152	6140	0.50	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.310	65	8031	0.53	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	=	106.00%
4) 1,2-Dichloroethane-d4	5.042	65	4634	0.53	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	106.00%
7) 1,2-Dichloropropane-d6	6.068	67	5658	0.54	ug/L	0.00
Spiked Amount	0.500	Range	60 - 140	Recovery	=	108.00%
8) Toluene-d8	7.317	98	15960	0.51	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	102.00%
10) 1,1,2,2-Tetrachloroeth...	10.216	84	2790	0.50	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	=	100.00%
Target Compounds						
					Qvalue	
3) Vinyl chloride	1.310	62	6829	0.45	ug/L	99
6) Trichloroethene	5.925	95	4749	0.46	ug/L	99
9) 1,2-Dibromoethane	8.368	107	2754	0.47	ug/L	100
12) 1,2,3-Trichloropropane	10.267	75	2456	0.48	ug/L	99
13) 1,2-Dibromo-3-chloropr...	12.426	75	298	0.42	ug/L	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV030321\
Data File : VV020438.D
Acq On : 03 Mar 2021 09:35
Operator : SY/MD
Sample : VSTDCCC0.5
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD0.5313

Quant Time: Mar 04 00:17:01 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM030221.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Mar 03 00:31:48 2021
Response via : Initial Calibration

