




Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV030323\
Data File : VV030140.D
Acq On : 03 Mar 2023 16:42
Operator : SY/MD
Sample : 01769-10
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK001

Manual IntegrationsAPPROVED

Quant Time: Mar 03 23:07:00 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR022323WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Mar 03 22:25:34 2023
Response via : Initial Calibration

Reviewed By :Krupa
Patel



TIC: VV030140.D\data.ms

(20) 2-Butanone-d5 (S)

3.847min (+ 0.013) 41.31 ug/L

response 67157

Ion	Exp%	Act%
46.10	100.00	100.00
77.10	12.60	13.57
0.00	0.00	0.00
0.00	0.00	0.00




Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV030323\
 Data File : VV030140.D
 Acq On : 03 Mar 2023 16:42
 Operator : SY/MD
 Sample : 01769-10
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK001

Manual IntegrationsAPPROVED

Quant Time: Mar 03 23:07:00 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR022323WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Mar 03 22:25:34 2023
 Response via : Initial Calibration

Reviewed By :Krupa
 Patel



(20) 2-Butanone-d5 (S)

3.847min (+ 0.013) 50.20 ug/L m

response 81601

Ion	Exp%	Act%
46.10	100.00	100.00
77.10	12.60	11.17
0.00	0.00	0.00
0.00	0.00	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV030323\
 Data File : VV030140.D
 Acq On : 03 Mar 2023 16:42
 Operator : SY/MD
 Sample : 01769-10
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK001

Manual IntegrationsAPPROVED

Quant Time: Mar 03 23:07:00 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR022323WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Mar 03 22:25:34 2023
 Response via : Initial Calibration

Reviewed By :Krupa
 Patel

Compound	R.T.	Q	Ion	Response	Conc	Units	Dev(Min)
-----03/06/2023							
Supervised By :Mahesh							
Padoda							
Internal Standards							
1) 1,4-Difluorobenzene	5.542	114		136258	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.793	117		135265	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.194	152		66192	5.000	ug/L	0.00
System Monitoring Compounds							
4) Vinyl Chloride-d3	1.282	65		36399	4.128	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	82.600%	
7) Chloroethane-d5	1.536	69		40052	4.839	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	96.800%	
11) 1,1-Dichloroethene-d2	2.063	63		61744	3.406	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	68.200%	
20) 2-Butanone-d5	3.847	46		81601m	50.200	ug/L	0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	100.400%	
24) Chloroform-d	4.262	84		91734	5.094	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	101.800%	
26) 1,2-Dichloroethane-d4	4.953	65		41042	5.582	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	111.600%	
32) Benzene-d6	4.970	84		159494	4.382	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	87.600%	
36) 1,2-Dichloropropane-d6	5.998	67		53745	4.908	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	98.200%	
41) Toluene-d8	7.252	98		133154	3.860	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	77.200%	
43) trans-1,3-Dichloroprop...	7.564	79		16173	4.317	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	86.400%	
46) 2-Hexanone-d5	8.037	63		76178	49.152	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	98.300%	
56) 1,1,2,2-Tetrachloroeth...	10.162	84		42405	5.444	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	108.800%	
66) 1,2-Dichlorobenzene-d4	11.571	152		54401	4.858	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	97.200%	
Target Compounds							
16) Methylene chloride	2.455	84		4782	0.428	ug/L	Qvalue 84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV030323\
Data File : VV030140.D
Acq On : 03 Mar 2023 16:42
Operator : SY/MD
Sample : 01769-10
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK001

Manual IntegrationsAPPROVED

Quant Time: Mar 03 23:07:00 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR022323WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Mar 03 22:25:34 2023
Response via : Initial Calibration

Reviewed By :Krupa
Patel

