



Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV030323\
Data File : VV030127.D
Acq On : 03 Mar 2023 08:48
Operator : SY/MD
Sample : BFB227
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BFB227

Integration File: rteint.p

Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR022323WMA.M
Title : TRACE VOA SFAM1.0
Last Update : Fri Mar 03 22:25:34 2023



AutoFind: Scans 2792, 2793, 2794; Background Corrected with Scan 2779

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	15.3	19856	PASS
75	95	30	80	47.8	62048	PASS
95	95	100	100	100.0	129904	PASS
96	95	5	9	6.5	8425	PASS
173	174	0.00	2	1.2	1362	PASS
174	95	50	120	89.9	116776	PASS
175	174	5	9	7.1	8327	PASS
176	174	95	105	95.1	111064	PASS
177	176	5	10	6.7	7396	PASS

m/z	Abundance
39.90	184.0
44.00	314.0

Instrument :
MSVOA_V
ClientSampleId :
BFB227

m/z	Abundance
36.00	945.0
37.00	4483.0
38.00	3835.0
39.10	1849.0
40.00	243.0
44.00	831.0
45.00	943.0
46.20	165.0
47.00	1757.0
48.00	605.0
49.10	4264.0
50.00	20632.0
51.00	6388.0
52.00	268.0
53.00	213.0
55.10	332.0
56.00	1716.0
57.00	3200.0
58.00	169.0
60.00	1177.0
61.00	5456.0
62.00	5789.0
63.00	4572.0
64.00	881.0
65.00	1059.0
67.10	280.0
68.00	12844.0
69.00	12655.0
70.00	943.0
72.00	635.0
73.00	5357.0
74.10	21480.0
75.10	64272.0
76.00	5539.0
77.00	787.0
78.00	587.0
78.90	4029.0
80.00	1135.0
80.90	4072.0
82.00	766.0
83.00	120.0
86.10	221.0
87.00	6267.0
88.00	5669.0
91.00	545.0
92.00	3848.0
93.00	5583.0
94.00	15747.0
95.00	133504.0
96.00	8695.0
97.00	194.0
103.90	656.0
104.80	188.0
105.90	606.0
106.80	139.0
107.90	105.0
112.80	130.0
114.90	142.0
115.90	522.0
116.90	960.0
117.90	493.0
118.90	790.0
127.90	521.0
128.90	221.0
129.80	425.0
130.90	182.0
134.90	261.0
136.80	226.0
141.00	1159.0
141.80	185.0
143.00	1695.0
143.90	177.0
144.90	635.0
146.00	222.0
147.80	316.0
148.90	141.0
149.80	115.0
152.90	139.0
155.00	292.0
156.90	278.0
158.70	142.0
160.90	144.0
170.90	243.0
171.90	1616.0
173.00	1408.0
174.00	117216.0
175.00	8613.0
176.00	111664.0
177.00	7479.0
177.90	195.0

Instrument :
MSVOA_V
ClientSampleId :
BFB227

m/z	Abundance
36.00	868.0
37.00	4248.0
38.00	3701.0
39.00	1752.0
39.90	269.0
44.00	877.0
45.00	901.0
45.90	139.0
46.20	134.0
47.00	1634.0
48.00	690.0
49.00	4309.0
50.00	20520.0
51.10	6361.0
52.00	229.0
53.00	212.0
55.00	344.0
56.00	1700.0
57.00	3154.0
58.00	165.0
60.00	1136.0
61.10	5516.0
62.00	5548.0
63.10	4636.0
64.00	820.0
65.00	1044.0
66.00	111.0
67.10	280.0
68.00	12676.0
69.00	12585.0
70.00	969.0
72.10	631.0
73.10	5449.0
74.10	21400.0
75.10	64144.0
76.00	5566.0
77.00	842.0
78.00	617.0
78.90	3989.0
80.00	1083.0
80.90	4109.0
81.90	838.0
82.90	123.0
85.80	186.0
87.00	6275.0
88.00	5751.0
90.90	511.0
92.00	3908.0
93.00	5659.0
94.00	15805.0
95.00	134336.0
96.00	8669.0
97.00	213.0
103.90	582.0
104.80	188.0
105.90	618.0
106.80	124.0
107.90	123.0
112.10	112.0
112.80	123.0
114.90	140.0
115.90	511.0
116.90	939.0
117.90	464.0
118.90	849.0
123.80	116.0
127.90	551.0
128.90	230.0
129.80	449.0
130.80	224.0
134.90	291.0
136.80	238.0
141.00	1226.0
141.80	168.0
143.00	1748.0
143.90	165.0
144.90	633.0
145.80	224.0
147.80	350.0
148.90	128.0
150.00	110.0
152.90	138.0
155.00	338.0
156.90	230.0
158.70	152.0
160.90	154.0
170.90	233.0
171.90	1682.0
173.00	1415.0
174.00	120880.0
175.00	8491.0
176.00	114696.0

Instrument :
MSVOA_V
ClientSampleId :
BFB227

177.007721.0

177.90159.0

Instrument :
MSVOA_V
ClientSampleId :
BFB227

m/z	Abundance
36.00	763.0
37.00	3705.0
38.00	3342.0
39.00	1562.0
39.90	279.0
44.00	860.0
45.00	775.0
47.00	1338.0
48.00	735.0
49.00	3977.0
50.00	18416.0
51.10	5683.0
52.00	203.0
52.90	188.0
55.00	317.0
56.00	1553.0
57.00	2847.0
58.00	152.0
60.00	1041.0
61.10	4957.0
62.00	4937.0
63.10	4275.0
64.00	615.0
65.00	955.0
66.10	127.0
67.00	284.0
68.00	11133.0
69.00	11157.0
70.00	958.0
72.10	592.0
73.10	5062.0
74.10	19064.0
75.10	57728.0
76.00	4938.0
77.00	779.0
77.90	597.0
78.90	3635.0
80.00	907.0
80.90	3809.0
81.90	834.0
82.80	130.0
85.80	164.0
87.00	5647.0
88.00	5038.0
90.90	469.0
92.00	3609.0
93.00	5219.0
94.00	14287.0
95.00	121872.0
96.00	7911.0
97.00	215.0
103.90	455.0
104.90	156.0
105.90	530.0
107.00	114.0
107.90	130.0
111.90	127.0
112.70	113.0
114.80	111.0
115.80	447.0
116.90	836.0
117.90	418.0
118.90	763.0
123.80	112.0
127.90	469.0
128.90	236.0
129.80	444.0
130.80	234.0
134.90	237.0
136.80	250.0
140.90	1224.0
141.90	163.0
143.00	1656.0
143.90	145.0
144.90	521.0
145.80	219.0
147.80	344.0
150.00	109.0
152.90	114.0
155.00	323.0
157.00	175.0
158.70	142.0
160.90	143.0
172.00	1699.0
173.10	1262.0
174.00	112232.0
175.00	7876.0
176.00	106832.0
177.00	6987.0
178.00	120.0

Instrument :
MSVOA_V
ClientSampleId :
BFB227