

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV030419\
 Data File : VV009479.D
 Acq On : 04 Mar 2019 18:34
 Operator : SY/MD
 Sample : K1660-14
 Misc : 25ML/MSVOA V/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK01

Quant Time: Mar 05 04:49:44 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR030419WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Mar 05 04:03:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	84511	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	83932	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	39243	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	28902	5.03	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	100.60%
7) Chloroethane-d5	1.58	69	26167	5.17	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	103.40%
11) 1,1-Dichloroethene-d2	2.13	63	55364	3.91	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	78.20%
20) 2-Butanone-d5	3.97	46	55561	55.18	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	110.36%
24) Chloroform-d	4.40	84	49826	4.96	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.20%
26) 1,2-Dichloroethane-d4	5.09	65	25909	5.23	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	104.60%
32) Benzene-d6	5.10	84	104834	5.14	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	102.80%
36) 1,2-Dichloropropane-d6	6.12	67	31912	5.25	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	105.00%
41) Toluene-d8	7.36	98	102450	4.99	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	99.80%
43) trans-1,3-Dichloropropene-	7.67	79	11766	4.51	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	90.20%
46) 2-Hexanone-d5	8.14	63	45168	51.90	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	103.80%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	22016	4.73	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	94.60%
64) 1,2-Dichlorobenzene-d4	11.68	152	36242	5.15	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	103.00%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV030419\
Data File : VV009479.D
Acq On : 04 Mar 2019 18:34
Operator : SY/MD
Sample : K1660-14
Misc : 25ML/MSVOA V/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK01

Quant Time: Mar 05 04:49:44 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR030419WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Mar 05 04:03:59 2019
Response via : Initial Calibration

