

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV030419\
 Data File : VV009472.D
 Acq On : 04 Mar 2019 15:19
 Operator : SY/MD
 Sample : VV0304WBL01
 Misc : 25ML/MSVOA V/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK67

Quant Time: Mar 05 04:19:25 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR030419WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Mar 05 04:03:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	94716	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	91449	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	42345	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	34383	5.34	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	106.80%
7) Chloroethane-d5	1.58	69	28344	4.99	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	99.80%
11) 1,1-Dichloroethene-d2	2.13	63	63366	4.00	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	80.00%
20) 2-Butanone-d5	3.98	46	59508	52.73	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	105.46%
24) Chloroform-d	4.41	84	55622	4.94	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.80%
26) 1,2-Dichloroethane-d4	5.09	65	29956	5.40	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	108.00%
32) Benzene-d6	5.10	84	117681	5.29	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	105.80%
36) 1,2-Dichloropropane-d6	6.12	67	34186	5.16	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	103.20%
41) Toluene-d8	7.36	98	116823	5.23	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	104.60%
43) trans-1,3-Dichloropropene-	7.67	79	14015	4.93	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	98.60%
46) 2-Hexanone-d5	8.14	63	49043	51.72	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	103.44%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	24474	4.83	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	96.60%
64) 1,2-Dichlorobenzene-d4	11.68	152	39673	5.22	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	104.40%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV030419\
Data File : VV009472.D
Acq On : 04 Mar 2019 15:19
Operator : SY/MD
Sample : VV0304WBL01
Misc : 25ML/MSVOA V/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK67

Quant Time: Mar 05 04:19:25 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR030419WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Mar 05 04:03:59 2019
Response via : Initial Calibration

