

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV030420\
 Data File : VV014731.D
 Acq On : 04 Mar 2020 13:53
 Operator : SY/MD
 Sample : L1677-18
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0CD6

Quant Time: Mar 05 06:47:56 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR022820WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Mar 05 06:28:00 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	331738	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	309220	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	153457	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	67687	2.61	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	52.20%
7) Chloroethane-d5	1.58	69	56872	2.80	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	56.00%#
11) 1,1-Dichloroethene-d2	2.13	63	88989	2.01	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	40.20%#
20) 2-Butanone-d5	3.96	46	161485	30.08	ug/L	-0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	60.16%
24) Chloroform-d	4.39	84	132337	2.84	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	56.80%#
26) 1,2-Dichloroethane-d4	5.08	65	69814	3.05	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	61.00%#
32) Benzene-d6	5.09	84	275029	3.07	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	61.40%#
36) 1,2-Dichloropropane-d6	6.11	67	88303	3.15	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	63.00%
41) Toluene-d8	7.35	98	241786	2.92	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	58.40%#
43) trans-1,3-Dichloropropene-	7.66	79	30914	2.85	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	57.00%
46) 2-Hexanone-d5	8.13	63	109425	24.93	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	49.86%
57) 1,1,2,2-Tetrachloroethane-	10.25	84	59864	2.98	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	59.60%#
64) 1,2-Dichlorobenzene-d4	11.67	152	87930	3.09	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	61.80%#

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV030420\
Data File : VV014731.D
Acq On : 04 Mar 2020 13:53
Operator : SY/MD
Sample : L1677-18
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0CD6

Quant Time: Mar 05 06:47:56 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR022820WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Mar 05 06:28:00 2020
Response via : Initial Calibration

