

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV030519\
Quantitation Report (QT Reviewed)


Data File : VV009501.D
Acq On : 05 Mar 2019 22:19
Operator : SY/MD
Sample : K1720-02
Misc : 2.90G/10ML/MSVOA_V/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BF8A8

Manual Integrations
APPROVED

MMDadoda
3/8/2019 12:10:44 AM

Quant Time: Mar 06 06:06:48 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOM2VLM030519S.M
Quant Title : VOC Analysis
QLast Update : Wed Mar 06 04:12:03 2019
Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV030519\
Quantitation Report (Qedit)




Data File : VV009501.D
Acq On : 05 Mar 2019 22:19
Operator : SY/MD
Sample : K1720-02
Misc : 2.90G/10ML/MSVOA_V/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BF8A8

Manual Integrations
APPROVED

MMDadoda
3/8/2019 12:10:44 AM

Quant Time: Mar 06 04:16:52 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOM2VLM030519S.M
Quant Title : VOC Analysis
QLast Update : Wed Mar 06 04:12:03 2019
Response via : Initial Calibration



(21) 2-Butanone

4.013min (-0.019) 0.13ug/L

response 98

Ion	Exp%	Act%
43.00	100	100
72.00	27.30	230.61#
0.00	0.00	0.00
0.00	0.00	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV030519\
Quantitation Report (Qedit)




Data File : VV009501.D
Acq On : 05 Mar 2019 22:19
Operator : SY/MD
Sample : K1720-02
Misc : 2.90G/10ML/MSVOA_V/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BF8A8

Manual Integrations
APPROVED

MMDadoda
3/8/2019 12:10:44 AM

Quant Time: Mar 06 04:16:52 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOM2VLM030519S.M
Quant Title : VOC Analysis
QLast Update : Wed Mar 06 04:12:03 2019
Response via : Initial Calibration



(21) 2-Butanone

4.042min (+0.010) 2.37ug/L m

203/08/19 SY

response 1839

Ion	Exp%	Act%
43.00	100	100
72.00	27.30	12.29#
0.00	0.00	0.00
0.00	0.00	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV030519\
Quantitation Report (QT Reviewed)

Data File : VV009501.D
Acq On : 05 Mar 2019 22:19
Operator : SY/MD
Sample : K1720-02
Misc : 2.90G/10ML/MSVOA_V/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BF8A8

Manual Integrations
APPROVED

MMDadoda
3/8/2019 12:10:44 AM

Quant Time: Mar 06 06:06:48 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOM2VLM030519S.M
Quant Title : VOC Analysis
QLast Update : Wed Mar 06 04:12:03 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	147618	25.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	129326	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	47320	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	38126	21.46	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	85.84%
7) Chloroethane-d5	1.58	69	66537	25.89	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	103.56%
10) 1,1-Dichloroethene-d2	2.13	63	77095	16.67	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	66.68%
20) 2-Butanone-d5	3.95	46	21805	36.01	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	72.02%
24) Chloroform-d	4.40	84	83929	23.87	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	95.48%
26) 1,2-Dichloroethane-d4	5.08	65	49594	24.79	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	99.16%
29) Benzene-d6	5.10	84	170418	26.61	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	106.44%
33) 1,2-Dichloropropane-d6	6.12	67	52070	27.55	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	110.20%
37) Toluene-d8	7.36	98	158080	25.32	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	101.28%
38) trans-1,3-Dichloropropene-	7.67	79	18406	20.09	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	80.36%
39) 2-Hexanone-d5	8.14	63	18514	41.97	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	83.94%
48) 1,1,2,2-Tetrachloroethane-	10.26	84	43517	21.60	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	86.40%
61) 1,2-Dichlorobenzene-d4	11.68	152	41930	24.27	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	97.08%

Target Compounds

					Qvalue
13) Acetone	2.21	43	6336	10.291	ug/L 89
16) Methylene chloride	2.54	84	5958	3.075	ug/L 93
21) 2-Butanone	4.04	43	1839m	2.367	ug/L

02/03/19 Sy

(#) = qualifier out of range (m) = manual integration (+) = signals summed