

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV030524\
 Data File : VV034500.D
 Acq On : 05 Mar 2024 11:38
 Operator : SY/MD
 Sample : VSTDICV0.5
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VICV206

Quant Time: Mar 06 04:21:19 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM030524.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Mar 06 04:17:25 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.566	114	11035	0.500	ug/L	0.00
5) Chlorobenzene-d5	8.800	117	10221	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.191	152	5432	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.280	65	6261	0.529	ug/L	0.00
Spiked Amount 0.500	Range 40 - 130		Recovery = 106.000%			
4) 1,2-Dichloroethane-d4	4.969	65	4140	0.565	ug/L	0.00
Spiked Amount 0.500	Range 70 - 130		Recovery = 112.000%			
7) 1,2-Dichloropropane-d6	5.995	67	4351	0.548	ug/L	0.00
Spiked Amount 0.500	Range 60 - 140		Recovery = 110.000%			
8) Toluene-d8	7.253	98	12570	0.534	ug/L	0.00
Spiked Amount 0.500	Range 70 - 130		Recovery = 106.000%			
10) 1,1,2,2-Tetrachloroeth...	10.146	84	3134	0.550	ug/L	0.00
Spiked Amount 0.500	Range 65 - 120		Recovery = 110.000%			
Target Compounds						
					Qvalue	
3) Vinyl chloride	1.280	62	6253	0.508	ug/L	98
6) Trichloroethene	5.852	95	4026	0.503	ug/L	99
9) 1,2-Dibromoethane	8.305	107	2599	0.536	ug/L	99
12) 1,2,3-Trichloropropane	10.197	75	2422	0.531	ug/L	98
13) 1,2-Dibromo-3-chloropr...	12.356	75	486	0.498	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV030524\
Data File : VV034500.D
Acq On : 05 Mar 2024 11:38
Operator : SY/MD
Sample : VSTDICV0.5
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VICV206

Quant Time: Mar 06 04:21:19 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM030524.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Mar 06 04:17:25 2024
Response via : Initial Calibration

