

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV030524\
 Data File : VV034495.D
 Acq On : 05 Mar 2024 09:23
 Operator : SY/MD
 Sample : VSTD0.0501
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VSTD0.05201

Quant Time: Mar 06 04:11:13 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM030524.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Mar 06 04:10:44 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.542	114	12209	0.500	ug/L	-0.02
5) Chlorobenzene-d5	8.800	117	11511	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.191	152	6017	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.280	65	604	0.040	ug/L	0.00
4) 1,2-Dichloroethane-d4	4.969	65	336	0.033	ug/L	0.00
7) 1,2-Dichloropropane-d6	5.995	67	398	0.037	ug/L	0.00
8) Toluene-d8	7.253	98	1193	0.039	ug/L	0.00
10) 1,1,2,2-Tetrachloroeth...	10.146	84	286	0.041	ug/L	0.00
Target Compounds						
						Qvalue
3) Vinyl chloride	1.280	62	704	0.048	ug/L #	82
6) Trichloroethene	5.852	95	444	0.046	ug/L	93
9) 1,2-Dibromoethane	8.305	107	245	0.038	ug/L #	88
12) 1,2,3-Trichloropropane	10.214	75	244	0.047	ug/L	99
13) 1,2-Dibromo-3-chloropr...	12.356	75	53	0.053	ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV030524\
Data File : VV034495.D
Acq On : 05 Mar 2024 09:23
Operator : SY/MD
Sample : VSTD0.0501
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD0.05201

Quant Time: Mar 06 04:11:13 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM030524.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Mar 06 04:10:44 2024
Response via : Initial Calibration

