

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV030524\
 Data File : VV034496.D
 Acq On : 05 Mar 2024 09:56
 Operator : SY/MD
 Sample : VSTD0.102
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VSTD0.1202

Quant Time: Mar 06 04:11:25 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM030524.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Mar 06 04:10:44 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.542	114	12256	0.500	ug/L	-0.02
5) Chlorobenzene-d5	8.800	117	11369	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.191	152	5995	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.280	65	1188	0.078	ug/L	0.00
4) 1,2-Dichloroethane-d4	4.969	65	710	0.070	ug/L	0.00
7) 1,2-Dichloropropane-d6	5.995	67	789	0.075	ug/L	0.00
8) Toluene-d8	7.253	98	2305	0.076	ug/L	0.00
10) 1,1,2,2-Tetrachloroeth...	10.146	84	550	0.080	ug/L	0.00
Target Compounds						
						Qvalue
3) Vinyl chloride	1.280	62	1284	0.088	ug/L	92
6) Trichloroethene	5.852	95	842	0.087	ug/L	95
9) 1,2-Dibromoethane	8.305	107	498	0.079	ug/L #	93
12) 1,2,3-Trichloropropane	10.214	75	461	0.089	ug/L	96
13) 1,2-Dibromo-3-chloropr...	12.356	75	108	0.109	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV030524\
Data File : VV034496.D
Acq On : 05 Mar 2024 09:56
Operator : SY/MD
Sample : VSTD0.102
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD0.1202

Quant Time: Mar 06 04:11:25 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM030524.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Mar 06 04:10:44 2024
Response via : Initial Calibration

