

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV030624\
 Data File : VV034506.D
 Acq On : 06 Mar 2024 09:30
 Operator : SY/MD
 Sample : VV0306WBL01
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK202

Quant Time: Mar 07 04:21:16 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM030524.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Mar 06 04:17:25 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.542	114	11300	0.500	ug/L	-0.02
5) Chlorobenzene-d5	8.800	117	10311	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.191	152	5243	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.280	65	6204	0.512	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	= 102.000%	
4) 1,2-Dichloroethane-d4	4.946	65	4099	0.546	ug/L	-0.02
Spiked Amount	0.500	Range	70 - 130	Recovery	= 110.000%	
7) 1,2-Dichloropropane-d6	5.996	67	4335	0.541	ug/L	0.00
Spiked Amount	0.500	Range	60 - 140	Recovery	= 108.000%	
8) Toluene-d8	7.254	98	12211	0.514	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	= 102.000%	
10) 1,1,2,2-Tetrachloroeth...	10.146	84	2995	0.521	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	= 104.000%	

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV030624\
Data File : VV034506.D
Acq On : 06 Mar 2024 09:30
Operator : SY/MD
Sample : VV0306WBL01
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK202

Quant Time: Mar 07 04:21:16 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM030524.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Mar 06 04:17:25 2024
Response via : Initial Calibration

