

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV030624\
 Data File : VV034508.D
 Acq On : 06 Mar 2024 10:34
 Operator : SY/MD
 Sample : VSTDCCC0.5EC
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VSTD0.5283

Quant Time: Mar 07 04:21:35 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM030524.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Mar 06 04:17:25 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.542	114	10971	0.500	ug/L	-0.02
5) Chlorobenzene-d5	8.800	117	10073	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.191	152	5041	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.280	65	6370	0.542	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	= 108.000%	
4) 1,2-Dichloroethane-d4	4.969	65	4127	0.566	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	= 114.000%	
7) 1,2-Dichloropropane-d6	5.995	67	4301	0.549	ug/L	0.00
Spiked Amount	0.500	Range	60 - 140	Recovery	= 110.000%	
8) Toluene-d8	7.253	98	12295	0.530	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	= 106.000%	
10) 1,1,2,2-Tetrachloroeth...	10.146	84	2950	0.525	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	= 106.000%	
Target Compounds						
					Qvalue	
3) Vinyl chloride	1.280	62	6330	0.517	ug/L	98
6) Trichloroethene	5.852	95	4060	0.515	ug/L	99
9) 1,2-Dibromoethane	8.305	107	2548	0.533	ug/L	99
12) 1,2,3-Trichloropropane	10.197	75	2309	0.546	ug/L	99
13) 1,2-Dibromo-3-chloropr...	12.356	75	487	0.538	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV030624\
Data File : VV034508.D
Acq On : 06 Mar 2024 10:34
Operator : SY/MD
Sample : VSTDCCC0.5EC
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD0.5283

Quant Time: Mar 07 04:21:35 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM030524.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Mar 06 04:17:25 2024
Response via : Initial Calibration

