

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV030719\
 Data File : VV009527.D
 Acq On : 07 Mar 2019 14:18
 Operator : SY/MD
 Sample : K1789-09
 Misc : 5.90G/10ML/MSVOA V/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BF0P5

Quant Time: Mar 08 04:29:05 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOM2VLM030519S.M
 Quant Title : VOC Analysis
 QLast Update : Fri Mar 08 03:01:51 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	172822	25.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	166794	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	74651	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	39689	19.08	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	76.32%
7) Chloroethane-d5	1.58	69	69258	23.02	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	92.08%
10) 1,1-Dichloroethene-d2	2.13	63	84541	15.62	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	62.48%
20) 2-Butanone-d5	3.95	46	21995	31.03	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	62.06%
24) Chloroform-d	4.40	84	91115	22.13	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	88.52%
26) 1,2-Dichloroethane-d4	5.08	65	51217	21.87	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	87.48%
29) Benzene-d6	5.10	84	184403	22.32	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	89.28%
33) 1,2-Dichloropropane-d6	6.12	67	56546	23.20	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	92.80%
37) Toluene-d8	7.36	98	179203	22.25	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	89.00%
38) trans-1,3-Dichloropropene-	7.67	79	23296	19.72	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	78.88%
39) 2-Hexanone-d5	8.14	63	17854	31.38	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	62.76%
48) 1,1,2,2-Tetrachloroethane-	10.26	84	47639	18.34	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	73.36%
61) 1,2-Dichlorobenzene-d4	11.67	152	61287	22.49	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	89.96%

Target Compounds

					Ovalue
16) Methylene chloride	2.54	84	4075	1.797 ug/L	94
22) cis-1,2-Dichloroethene	3.96	96	4177	1.793 ug/L	93
35) Trichloroethene	5.96	95	2626	1.060 ug/L	94
47) Tetrachloroethene	8.02	164	142381	75.790 ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV030719\
Data File : VV009527.D
Acq On : 07 Mar 2019 14:18
Operator : SY/MD
Sample : K1789-09
Misc : 5.90G/10ML/MSVOA V/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BF0P5

Quant Time: Mar 08 04:29:05 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOM2VLM030519S.M
Quant Title : VOC Analysis
QLast Update : Fri Mar 08 03:01:51 2019
Response via : Initial Calibration

