

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV030719\
 Data File : VV009537.D
 Acq On : 07 Mar 2019 18:45
 Operator : SY/MD
 Sample : K1789-06
 Misc : 5.83G/10ML/MSVOA V/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BF0P2

Quant Time: Mar 08 05:00:31 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOM2VLM030519S.M
 Quant Title : VOC Analysis
 QLast Update : Fri Mar 08 03:01:51 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	128820	25.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	125345	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	57558	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	31470	20.29	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	81.16%
7) Chloroethane-d5	1.58	69	52460	23.39	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	93.56%
10) 1,1-Dichloroethene-d2	2.13	63	67990	16.85	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	67.40%
20) 2-Butanone-d5	3.95	46	24506	46.38	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	92.76%
24) Chloroform-d	4.40	84	73451	23.93	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	95.72%
26) 1,2-Dichloroethane-d4	5.08	65	43426	24.87	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	99.48%
29) Benzene-d6	5.10	84	146770	23.64	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	94.56%
33) 1,2-Dichloropropane-d6	6.12	67	46263	25.25	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	101.00%
37) Toluene-d8	7.36	98	139588	23.06	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	92.24%
38) trans-1,3-Dichloropropene-	7.67	79	19055	21.46	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	85.84%
39) 2-Hexanone-d5	8.14	63	18173	42.50	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	85.00%
48) 1,1,2,2-Tetrachloroethane-	10.26	84	44951	23.02	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	92.08%
61) 1,2-Dichlorobenzene-d4	11.67	152	50101	23.84	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	95.36%

Target Compounds

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
16) Methylene chloride	2.53	84	2792	1.651	ug/L	92
22) cis-1,2-Dichloroethene	3.96	96	42853	24.674	ug/L	98
35) Trichloroethene	5.96	95	48691	26.159	ug/L	98
47) Tetrachloroethene	8.03	164	2414290	1710.098	ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV030719\
Data File : VV009537.D
Acq On : 07 Mar 2019 18:45
Operator : SY/MD
Sample : K1789-06
Misc : 5.83G/10ML/MSVOA V/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BF0P2

Quant Time: Mar 08 05:00:31 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOM2VLM030519S.M
Quant Title : VOC Analysis
QLast Update : Fri Mar 08 03:01:51 2019
Response via : Initial Calibration

