

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV030823\
 Data File : VV030145.D
 Acq On : 08 Mar 2023 12:23
 Operator : SY/MD
 Sample : 01838-03
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C1170

Quant Time: Mar 09 00:36:40 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR022323WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Mar 03 22:25:34 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.542	114	142923	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.792	117	140971	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.194	152	70861	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.281	65	34414	3.721	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	74.400%
7) Chloroethane-d5	1.535	69	38687	4.457	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	89.200%
11) 1,1-Dichloroethene-d2	2.063	63	55200	2.903	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	58.000%#
20) 2-Butanone-d5	3.815	46	89238	52.338	ug/L	-0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	104.680%
24) Chloroform-d	4.259	84	90963	4.815	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.400%
26) 1,2-Dichloroethane-d4	4.953	65	38921	5.047	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	101.000%
32) Benzene-d6	4.969	84	156012	4.113	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	82.200%
36) 1,2-Dichloropropane-d6	5.998	67	53594	4.696	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	94.000%
41) Toluene-d8	7.252	98	127714	3.552	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	71.000%
43) trans-1,3-Dichloroprop...	7.561	79	17111	4.383	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	87.600%
46) 2-Hexanone-d5	8.034	63	75920	47.002	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	94.000%
56) 1,1,2,2-Tetrachloroeth...	10.162	84	41797	5.148	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	103.000%
66) 1,2-Dichlorobenzene-d4	11.570	152	58809	4.905	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	98.200%
Target Compounds						Qvalue
13) Acetone	2.140	43	5147	4.084	ug/L	# 55
16) Methylene chloride	2.455	84	4606	0.393	ug/L	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV030823\
Data File : VV030145.D
Acq On : 08 Mar 2023 12:23
Operator : SY/MD
Sample : 01838-03
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C1170

Quant Time: Mar 09 00:36:40 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR022323WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Mar 03 22:25:34 2023
Response via : Initial Calibration

