

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV030824\
 Data File : VV034523.D
 Acq On : 08 Mar 2024 11:39
 Operator : SY/MD
 Sample : P1679-01
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 E0700

Quant Time: Mar 08 23:19:50 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM022924WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Mar 08 23:17:44 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	320054	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.783	117	302888	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	144917	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	111722	40.267	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	80.540%
7) Chloroethane-d5	1.529	69	99509	46.036	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	92.080%
11) 1,1-Dichloroethene-d2	2.059	65	51150	40.764	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	81.520%
21) 2-Butanone-d5	3.780	46	193318	104.190	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	104.190%
24) Chloroform-d	4.249	84	229233	46.728	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.460%
26) 1,2-Dichloroethane-d4	4.940	65	158901	49.568	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.140%
32) Benzene-d6	4.960	84	430406	47.084	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.160%
36) 1,2-Dichloropropane-d6	5.989	67	136434	48.906	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	97.820%
41) Toluene-d8	7.243	98	377545	45.639	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	91.280%
43) trans-1,3-Dichloroprop...	7.554	79	62234	44.424	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	88.840%
47) 2-Hexanone-d5	8.021	63	143540	103.039	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	103.040%
56) 1,1,2,2-Tetrachloroeth...	10.149	84	175405	47.778	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	95.560%
66) 1,2-Dichlorobenzene-d4	11.558	152	149285	50.398	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	100.800%
Target Compounds						
25) Chloroform	4.281	83	29462	6.256	ug/L	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV030824\
Data File : VV034523.D
Acq On : 08 Mar 2024 11:39
Operator : SY/MD
Sample : P1679-01
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
E0700

Quant Time: Mar 08 23:19:50 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM022924WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Mar 08 23:17:44 2024
Response via : Initial Calibration

