

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV030921\
 Data File : VV020475.D
 Acq On : 09 Mar 2021 16:54
 Operator : SY/MD
 Sample : M1617-01
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BG0M7

Manual Integrations
 APPROVED

MMDadoda
 3/10/2021 1:42:17 PM

Quant Time: Mar 10 00:20:49 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SOMVTR022221WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Mar 10 00:20:01 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	301559	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.857	117	298514	5.00	ug/L	0.00
61) 1,4-Dichlorobenzene-d4	11.255	152	143018	5.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	93306	4.39	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	87.80%
7) Chloroethane-d5	1.568	69	74651	4.47	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	89.40%
11) 1,1-Dichloroethene-d2	2.108	63	124854	3.02	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	60.40%
20) 2-Butanone-d5	3.918	46	144344	39.24	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	78.48%
24) Chloroform-d	4.349	84	159226	4.19	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	83.80%
26) 1,2-Dichloroethane-d4	5.034	65	77249	4.18	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	83.60%
32) Benzene-d6	5.050	84	326912	4.37	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	87.40%
36) 1,2-Dichloropropane-d6	6.069	67	92304	4.17	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	83.40%
41) Toluene-d8	7.317	98	301606	4.18	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	83.60%
45) trans-1,3-Dichloroprop...	7.625	79	32930	3.89	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	77.80%
48) 2-Hexanone-d5	8.095	63	89408	30.14	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	60.28%
59) 1,1,2,2-Tetrachloroeth...	10.223	84	49262	3.13	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	62.60%#
65) 1,2-Dichlorobenzene-d4	11.632	152	120314	4.74	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	94.80%
Target Compounds					Qvalue	
13) Acetone	2.214	43	6686m	2.00	ug/L	
16) Methylene chloride	2.507	84	5734	0.21	ug/L	100
25) Chloroform	4.378	83	7215	0.14	ug/L	98
50) 2-Hexanone	8.146	43	20023	2.23	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV030921\
Data File : VV020475.D
Acq On : 09 Mar 2021 16:54
Operator : SY/MD
Sample : M1617-01
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1


Instrument :
MSVOA_V
ClientSampleId :
BG0M7

Manual Integrations
APPROVED

MMDadoda
3/10/2021 1:42:17 PM

Quant Time: Mar 10 00:20:49 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SOMVTR022221WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Mar 10 00:20:01 2021
Response via : Initial Calibration









#13
Acetone
Concen: 2.00 ug/L m
RT: 2.214 min Scan# 365
Delta R.T. -0.003 min
Lab File: VV020475.D
Acq: 09 Mar 2021 16:54

Tgt Ion: 43 Resp: 6686
Ion Ratio Lower Upper
43 100
58 1.4 0.0 18.6

Instrument :
MSVOA_V
ClientSampleId :
BG0M7




Manual Integrations
APPROVED


MMDadoda
3/10/2021 1:42:17 PM



#16
Methylene chloride
Concen: 0.21 ug/L
RT: 2.507 min Scan# 456
Delta R.T. 0.003 min
Lab File: VV020475.D
Acq: 09 Mar 2021 16:54

Tgt Ion: 84 Resp: 5734
Ion Ratio Lower Upper
84 100
86 65.9 46.5 86.3
49 113.2 79.2 147.2





#25





Chloroform
Concen: 0.14 ug/L
RT: 4.378 min Scan# 1038
Delta R.T. 0.003 min
Lab File: VV020475.D
Acq: 09 Mar 2021 16:54

Instrument :
MSVOA_V
ClientSampleId :
BG0M7

Tgt Ion: 83 Resp: 7215
Ion Ratio Lower Upper
83 100
85 69.0 47.3 87.9

Manual Integrations
APPROVED

MMDadoda
3/10/2021 1:42:17 PM



#50

2-Hexanone
Concen: 2.23 ug/L
RT: 8.146 min Scan# 2210
Delta R.T. 0.003 min
Lab File: VV020475.D
Acq: 09 Mar 2021 16:54

Tgt Ion: 43 Resp: 20023
Ion Ratio Lower Upper
43 100
58 53.7 41.6 62.4
57 19.2 15.4 23.0
100 14.1 11.2 16.8

