

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW030922\
 Data File : VW024887.D
 Acq On : 09 Mar 2022 17:53
 Operator : SY/MD
 Sample : N1856-01
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0D49

Quant Time: Mar 10 05:50:35 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM022822WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Mar 10 05:49:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	302729	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	291182	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.246	152	132451	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	110838	36.008	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	72.020%
7) Chloroethane-d5	1.568	69	97604	40.415	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	80.840%
11) 1,1-Dichloroethene-d2	2.108	63	155939	30.183	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	60.360%
21) 2-Butanone-d5	3.892	46	172327	98.372	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	98.370%
24) Chloroform-d	4.349	84	226759	42.200	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	84.400%
26) 1,2-Dichloroethane-d4	5.031	65	139657	44.050	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.100%
32) Benzene-d6	5.050	84	470113	43.558	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	87.120%
36) 1,2-Dichloropropane-d6	6.069	67	152112	45.236	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	90.480%
41) Toluene-d8	7.313	98	407412	42.609	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	85.220%
43) trans-1,3-Dichloroprop...	7.622	79	61097	41.815	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	83.640%
47) 2-Hexanone-d5	8.088	63	123293	96.597	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	96.600%
56) 1,1,2,2-Tetrachloroeth...	10.214	84	184822	46.340	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	92.680%
66) 1,2-Dichlorobenzene-d4	11.622	152	143322	46.975	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	93.940%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV030922\
Data File : VV024887.D
Acq On : 09 Mar 2022 17:53
Operator : SY/MD
Sample : N1856-01
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0D49

Quant Time: Mar 10 05:50:35 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM022822WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Mar 10 05:49:38 2022
Response via : Initial Calibration

