

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW030922\
 Data File : VW024894.D
 Acq On : 10 Mar 2022 00:08
 Operator : SY/MD
 Sample : N1856-09
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0D61

Quant Time: Mar 10 05:51:58 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM022822WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Mar 10 05:49:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	299931	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	290085	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.246	152	130539	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	109411	35.876	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	71.760%
7) Chloroethane-d5	1.568	69	98894	41.331	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	82.660%
11) 1,1-Dichloroethene-d2	2.108	63	156270	30.529	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	61.060%
21) 2-Butanone-d5	3.899	46	167826	96.696	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	96.700%
24) Chloroform-d	4.349	84	226537	42.552	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	85.100%
26) 1,2-Dichloroethane-d4	5.031	65	140289	44.662	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	89.320%
32) Benzene-d6	5.050	84	471666	43.867	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	87.740%
36) 1,2-Dichloropropane-d6	6.069	67	151138	45.117	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	90.240%
41) Toluene-d8	7.313	98	408798	42.916	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	85.840%
43) trans-1,3-Dichloroprop...	7.622	79	59672	40.995	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	81.980%
47) 2-Hexanone-d5	8.088	63	119753	94.179	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	94.180%
56) 1,1,2,2-Tetrachloroeth...	10.214	84	186890	47.036	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	94.080%
66) 1,2-Dichlorobenzene-d4	11.622	152	141437	47.036	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	94.080%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV030922\
Data File : VV024894.D
Acq On : 10 Mar 2022 00:08
Operator : SY/MD
Sample : N1856-09
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0D61

Quant Time: Mar 10 05:51:58 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM022822WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Mar 10 05:49:38 2022
Response via : Initial Calibration

