

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\Data\VW030922\  
 Data File : VW024897.D  
 Acq On : 10 Mar 2022 01:18  
 Operator : SY/MD  
 Sample : N1855-03  
 Misc : 5.0mL/MSVOA\_V/WATER  
 ALS Vial : 16 Sample Multiplier: 1

Instrument :  
 MSVOA\_V  
 ClientSampleId :  
 C0D25

Quant Time: Mar 10 05:52:34 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\SFAMVLM022822WMA.M  
 Quant Title : VOC Analysis  
 QLast Update : Thu Mar 10 05:49:38 2022  
 Response via : Initial Calibration

| Compound                      | R.T.    | QIon  | Response | Conc     | Units | Dev(Min) |
|-------------------------------|---------|-------|----------|----------|-------|----------|
| -----                         |         |       |          |          |       |          |
| Internal Standards            |         |       |          |          |       |          |
| 1) 1,4-Difluorobenzene        | 5.616   | 114   | 298453   | 50.000   | ug/L  | 0.00     |
| 28) Chlorobenzene-d5          | 8.850   | 117   | 293396   | 50.000   | ug/L  | 0.00     |
| 58) 1,4-Dichlorobenzene-d4    | 11.249  | 152   | 132450   | 50.000   | ug/L  | 0.00     |
| System Monitoring Compounds   |         |       |          |          |       |          |
| 4) Vinyl Chloride-d3          | 1.307   | 65    | 105070   | 34.623   | ug/L  | 0.00     |
| Spiked Amount                 | 50.000  | Range | 60 - 135 | Recovery | =     | 69.240%  |
| 7) Chloroethane-d5            | 1.568   | 69    | 94415    | 39.655   | ug/L  | 0.00     |
| Spiked Amount                 | 50.000  | Range | 70 - 130 | Recovery | =     | 79.300%  |
| 11) 1,1-Dichloroethene-d2     | 2.108   | 63    | 149220   | 29.296   | ug/L  | 0.00     |
| Spiked Amount                 | 50.000  | Range | 60 - 125 | Recovery | =     | 58.600%# |
| 21) 2-Butanone-d5             | 3.896   | 46    | 160500   | 92.933   | ug/L  | 0.00     |
| Spiked Amount                 | 100.000 | Range | 40 - 130 | Recovery | =     | 92.930%  |
| 24) Chloroform-d              | 4.349   | 84    | 219649   | 41.463   | ug/L  | 0.00     |
| Spiked Amount                 | 50.000  | Range | 70 - 125 | Recovery | =     | 82.920%  |
| 26) 1,2-Dichloroethane-d4     | 5.031   | 65    | 136977   | 43.823   | ug/L  | 0.00     |
| Spiked Amount                 | 50.000  | Range | 70 - 125 | Recovery | =     | 87.640%  |
| 32) Benzene-d6                | 5.050   | 84    | 452940   | 41.650   | ug/L  | 0.00     |
| Spiked Amount                 | 50.000  | Range | 70 - 125 | Recovery | =     | 83.300%  |
| 36) 1,2-Dichloropropane-d6    | 6.069   | 67    | 146900   | 43.357   | ug/L  | 0.00     |
| Spiked Amount                 | 50.000  | Range | 70 - 120 | Recovery | =     | 86.720%  |
| 41) Toluene-d8                | 7.314   | 98    | 395149   | 41.015   | ug/L  | 0.00     |
| Spiked Amount                 | 50.000  | Range | 80 - 120 | Recovery | =     | 82.020%  |
| 43) trans-1,3-Dichloroprop... | 7.622   | 79    | 58041    | 39.424   | ug/L  | 0.00     |
| Spiked Amount                 | 50.000  | Range | 60 - 125 | Recovery | =     | 78.840%  |
| 47) 2-Hexanone-d5             | 8.088   | 63    | 112966   | 87.838   | ug/L  | 0.00     |
| Spiked Amount                 | 100.000 | Range | 45 - 130 | Recovery | =     | 87.840%  |
| 56) 1,1,2,2-Tetrachloroeth... | 10.214  | 84    | 179206   | 44.593   | ug/L  | 0.00     |
| Spiked Amount                 | 50.000  | Range | 65 - 120 | Recovery | =     | 89.180%  |
| 66) 1,2-Dichlorobenzene-d4    | 11.622  | 152   | 139223   | 45.632   | ug/L  | 0.00     |
| Spiked Amount                 | 50.000  | Range | 80 - 120 | Recovery | =     | 91.260%  |

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\Data\VV030922\  
Data File : VV024897.D  
Acq On : 10 Mar 2022 01:18  
Operator : SY/MD  
Sample : N1855-03  
Misc : 5.0mL/MSVOA\_V/WATER  
ALS Vial : 16 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
C0D25

Quant Time: Mar 10 05:52:34 2022  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\SFAMVLM022822WMA.M  
Quant Title : VOC Analysis  
QLast Update : Thu Mar 10 05:49:38 2022  
Response via : Initial Calibration

