

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW030922\
 Data File : VW024898.D
 Acq On : 10 Mar 2022 01:42
 Operator : SY/MD
 Sample : N1855-04
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0D26

Quant Time: Mar 10 05:52:44 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM022822WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Mar 10 05:49:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	302157	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	293297	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.246	152	132259	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	101833	33.145	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	66.280%
7) Chloroethane-d5	1.571	69	94015	39.003	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	78.000%
11) 1,1-Dichloroethene-d2	2.108	63	147621	28.627	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	57.260%#
21) 2-Butanone-d5	3.896	46	160088	91.558	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	91.560%
24) Chloroform-d	4.349	84	217784	40.606	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	81.220%
26) 1,2-Dichloroethane-d4	5.031	65	135547	42.834	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	85.660%
32) Benzene-d6	5.050	84	447305	41.146	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	82.300%
36) 1,2-Dichloropropane-d6	6.069	67	145595	42.986	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	85.980%
41) Toluene-d8	7.313	98	389658	40.459	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	80.920%
43) trans-1,3-Dichloroprop...	7.622	79	56348	38.287	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	76.580%
47) 2-Hexanone-d5	8.088	63	117493	91.389	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	91.390%
56) 1,1,2,2-Tetrachloroeth...	10.214	84	179923	44.786	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	89.580%
66) 1,2-Dichlorobenzene-d4	11.622	152	136429	44.780	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	89.560%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV030922\
Data File : VV024898.D
Acq On : 10 Mar 2022 01:42
Operator : SY/MD
Sample : N1855-04
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0D26

Quant Time: Mar 10 05:52:44 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM022822WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Mar 10 05:49:38 2022
Response via : Initial Calibration

