

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW030922\
 Data File : VW024895.D
 Acq On : 10 Mar 2022 00:31
 Operator : SY/MD
 Sample : N1855-01
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0D22

Quant Time: Mar 10 05:52:10 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM022822WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Mar 10 05:49:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	303705	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	293668	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.246	152	132741	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	108513	35.139	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	70.280%
7) Chloroethane-d5	1.568	69	100541	41.497	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	83.000%
11) 1,1-Dichloroethene-d2	2.108	63	155143	29.932	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	59.860%#
21) 2-Butanone-d5	3.896	46	161730	92.026	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	92.030%
24) Chloroform-d	4.349	84	227639	42.228	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	84.460%
26) 1,2-Dichloroethane-d4	5.031	65	142085	44.671	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	89.340%
32) Benzene-d6	5.050	84	477240	43.844	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	87.680%
36) 1,2-Dichloropropane-d6	6.069	67	152598	44.997	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	90.000%
41) Toluene-d8	7.313	98	413209	42.850	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	85.700%
43) trans-1,3-Dichloroprop...	7.622	79	60949	41.361	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	82.720%
47) 2-Hexanone-d5	8.088	63	114546	88.984	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	88.980%
56) 1,1,2,2-Tetrachloroeth...	10.214	84	184952	45.980	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	91.960%
66) 1,2-Dichlorobenzene-d4	11.622	152	144096	47.125	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	94.260%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV030922\
Data File : VV024895.D
Acq On : 10 Mar 2022 00:31
Operator : SY/MD
Sample : N1855-01
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0D22

Quant Time: Mar 10 05:52:10 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM022822WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Mar 10 05:49:38 2022
Response via : Initial Calibration

