

Data Path : Z:\voasrv\HPCHEM1\MSVOA V\Data\VV031019\
 Data File : VV009562.D
 Acq On : 10 Mar 2019 12:22
 Operator : SY/MD
 Sample : VSTDCCC025
 Misc : 5.00G/10ML/MSVOA V/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 LabSampleId :
 VSTD02574

Quant Time: Mar 11 04:44:23 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOM2VLM030519S.M
 Quant Title : VOC Analysis
 QLast Update : Sat Mar 09 01:24:39 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 200%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Difluorobenzene	1.000	1.000	0.0	107	0.00
2 T	Dichlorodifluoromethane	0.288	0.294	-2.1	110	0.00
3 T	Chloromethane	0.307	0.289	5.9	102	0.00
4 S	Vinyl Chloride-d3	0.301	0.257	14.6	89	0.00
5 T	Vinyl chloride	0.338	0.324	4.1	105	0.00
6 T	Bromomethane	0.421	0.404	4.0	105	0.00
7 S	Chloroethane-d5	0.435	0.388	10.8	91	0.00
8 T	Chloroethane	0.376	0.283	24.7	89	0.00
9 T	Trichlorofluoromethane	0.694	0.695	-0.1	109	0.00
10 S	1,1-Dichloroethene-d2	0.783	0.714	8.8	99	0.00
11 T	1,1,2-Trichloro-1,2,2-trifl	0.388	0.398	-2.6	113	0.00
12 T	1,1-Dichloroethene	0.332	0.323	2.7	108	0.00
13 T	Acetone	0.104	0.075	27.9	84	0.00
14 T	Carbon disulfide	0.842	0.814	3.3	101	0.00
15 T	Methyl Acetate	0.169	0.126	25.4	81	0.00
16 T	Methylene chloride	0.328	0.287	12.5	105	0.00
17 T	Methyl tert-butyl Ether	0.744	0.672	9.7	101	0.00
18 T	trans-1,2-Dichloroethene	0.299	0.297	0.7	109	0.00
19 T	1,1-Dichloroethane	0.526	0.535	-1.7	112	0.00
20 S	2-Butanone-d5	0.103	0.082	20.4	87	0.00
21	2-Butanone	0.132	0.088	33.3#	73	0.00
22 T	cis-1,2-Dichloroethene	0.337	0.351	-4.2	114	0.00
23 T	Bromochloromethane	0.155	0.153	1.3	107	0.00
24 S	Chloroform-d	0.596	0.576	3.4	105	0.00
25 T	Chloroform	0.587	0.591	-0.7	110	0.00
26 S	1,2-Dichloroethane-d4	0.339	0.313	7.7	99	0.00
27 T	1,2-Dichloroethane	0.408	0.385	5.6	103	0.00
28 I	Chlorobenzene-d5	1.000	1.000	0.0	104	0.00
29 S	Benzene-d6	1.238	1.189	4.0	99	0.00
30 T	Cyclohexane	0.531	0.526	0.9	104	0.00
31 T	1,1,1-Trichloroethane	0.559	0.584	-4.5	109	0.00
32 T	Carbon tetrachloride	0.472	0.498	-5.5	113	0.00
33 S	1,2-Dichloropropane-d6	0.365	0.350	4.1	101	0.00
34 T	Benzene	1.292	1.329	-2.9	109	0.00
35 T	Trichloroethene	0.371	0.379	-2.2	109	0.00
36 T	Methylcyclohexane	0.605	0.597	1.3	105	0.00
37 S	Toluene-d8	1.207	1.181	2.2	101	0.00
38 S	trans-1,3-Dichloropropene-d	0.177	0.161	9.0	95	0.00
39 S	2-Hexanone-d5	0.085	0.068	20.0	84	0.00
40 T	1,2-Dichloropropane	0.333	0.323	3.0	108	0.00
41 T	Bromodichloromethane	0.450	0.464	-3.1	109	0.00
42 T	cis-1,3-Dichloropropene	0.517	0.517	0.0	102	0.00
43 T	4-Methyl-2-pentanone	0.263	0.204	22.4	86	0.00
44 T	Toluene	1.460	1.499	-2.7	108	0.00
45 T	trans-1,3-Dichloropropene	0.452	0.444	1.8	103	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA V\Data\VV031019\
 Data File : VV009562.D
 Acq On : 10 Mar 2019 12:22
 Operator : SY/MD
 Sample : VSTDCCC025
 Misc : 5.00G/10ML/MSVOA V/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 LabSampleId :
 VSTD02574

Quant Time: Mar 11 04:44:23 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOM2VLM030519S.M
 Quant Title : VOC Analysis
 QLast Update : Sat Mar 09 01:24:39 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 200%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
46 T	1,1,2-Trichloroethane	0.277	0.268	3.2	102	0.00
47 T	Tetrachloroethene	0.282	0.288	-2.1	107	0.00
48 S	1,1,2,2-Tetrachloroethane-d	0.389	0.349	10.3	96	0.00
49 T	2-Hexanone	0.199	0.151	24.1	87	0.00
50 T	Dibromochloromethane	0.335	0.328	2.1	105	0.00
51 T	1,2-Dibromoethane	0.286	0.265	7.3	98	0.00
52 T	Chlorobenzene	0.968	0.967	0.1	106	0.00
53 T	Ethylbenzene	1.671	1.707	-2.2	108	0.00
54 T	m,p-Xylene	0.632	0.644	-1.9	108	0.00
55 T	o-xylene	0.610	0.623	-2.1	105	0.00
56 T	Styrene	1.003	1.025	-2.2	108	0.00
57 T	Isopropylbenzene	1.682	1.714	-1.9	108	0.00
58 T	1,1,2,2-Tetrachloroethane	0.391	0.355	9.2	96	0.00
59	1,2,3-Trichloropropane	0.282	0.245	13.1	94	0.00
60 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	106	0.00
61 S	1,2-Dichlorobenzene-d4	0.913	0.863	5.5	101	0.00
62 T	Bromoform	0.406	0.370	8.9	97	0.00
63 T	1,3-Dichlorobenzene	1.562	1.550	0.8	107	0.00
64 T	1,4-Dichlorobenzene	1.596	1.535	3.8	106	0.00
65 T	1,2-Dichlorobenzene	1.456	1.428	1.9	104	0.00
66 T	1,2-Dibromo-3-chloropropane	0.146	0.118	19.2	80	0.00
67	1,3,5-Trichlorobenzene	1.120	1.114	0.5	107	0.00
68 T	1,2,4-trichlorobenzene	0.897	0.858	4.3	103	0.00
69	Naphthalene	1.889	1.621	14.2	92	0.00
70 T	1,2,3-Trichlorobenzene	0.825	0.805	2.4	105	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0