

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV031020\
 Data File : VV014799.D
 Acq On : 10 Mar 2020 13:44
 Operator : SY/MD
 Sample : L1858-05 50X
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BFRF6

Quant Time: Mar 11 03:22:45 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM030620WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Mar 11 01:57:25 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	405742	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	388771	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	189140	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	105601	38.46	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	76.92%
7) Chloroethane-d5	1.58	69	89979	41.45	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	82.90%
11) 1,1-Dichloroethene-d2	2.12	63	140738	30.85	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	61.70%
21) 2-Butanone-d5	3.94	46	182265	89.00	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	89.00%
24) Chloroform-d	4.38	84	237120	45.61	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.22%
26) 1,2-Dichloroethane-d4	5.06	65	161449	45.72	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.44%
32) Benzene-d6	5.08	84	486049	47.35	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.70%
36) 1,2-Dichloropropane-d6	6.10	67	152880	44.93	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	89.86%
41) Toluene-d8	7.34	98	460501	45.80	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	91.60%
43) trans-1,3-Dichloropropene-	7.64	79	71881	43.27	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	86.54%
47) 2-Hexanone-d5	8.12	63	146754	97.15	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	97.15%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	217334	46.44	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	92.88%
64) 1,2-Dichlorobenzene-d4	11.65	152	188399	50.26	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	100.52%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV031020\
Data File : VV014799.D
Acq On : 10 Mar 2020 13:44
Operator : SY/MD
Sample : L1858-05 50X
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BFRF6

Quant Time: Mar 11 03:22:45 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM030620WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Mar 11 01:57:25 2020
Response via : Initial Calibration

